

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042883.D
 Acq On : 06 Aug 2024 20:38
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 07 02:59:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	132749	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	243995	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83716	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105743	58.025	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.040%
35) Dibromofluoromethane	5.385	113	92497	57.979	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	115.960%
50) Toluene-d8	8.647	98	279575	49.364	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.720%
62) 4-Bromofluorobenzene	11.079	95	95948	47.145	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042883.D
Acq On : 06 Aug 2024 20:38
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 07 02:59:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

