

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042892.D
 Acq On : 07 Aug 2024 01:43
 Operator : JC/MD
 Sample : P3448-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP3

Quant Time: Aug 07 06:31:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	261160	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108874	57.899	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.800%
35) Dibromofluoromethane	5.385	113	91939	53.841	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.680%
50) Toluene-d8	8.647	98	320574	52.882	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.760%
62) 4-Bromofluorobenzene	11.079	95	116286	53.382	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.760%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	84052	103.054	ug/l	99
18) Methyl Acetate	2.715	43	5126	1.982	ug/l	98
20) Methylene Chloride	2.788	84	12913	6.796	ug/l #	86
25) 2-Butanone	4.586	43	11660	9.103	ug/l #	86
43) Isopropyl Acetate	6.348	43	157306	36.720	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042892.D
Acq On : 07 Aug 2024 01:43
Operator : JC/MD
Sample : P3448-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP3

Quant Time: Aug 07 06:31:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

