

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042896.D
 Acq On : 07 Aug 2024 03:16
 Operator : JC/MD
 Sample : P3448-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP7

Quant Time: Aug 07 06:33:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	146751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	283068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	279166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113380	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113589	56.383	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.760%	
35) Dibromofluoromethane	5.385	113	96785	52.293	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.580%	
50) Toluene-d8	8.647	98	355347	54.082	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.160%	
62) 4-Bromofluorobenzene	11.079	95	126909	53.750	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.500%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	80546	92.178	ug/l	95
18) Methyl Acetate	2.715	43	7058	2.547	ug/l	96
20) Methylene Chloride	2.782	84	13037	6.351	ug/l	94
25) 2-Butanone	4.580	43	9855	7.182	ug/l	100
43) Isopropyl Acetate	6.342	43	159778	34.411	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042896.D
Acq On : 07 Aug 2024 03:16
Operator : JC/MD
Sample : P3448-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP7

Quant Time: Aug 07 06:33:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

