

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042913.D
 Acq On : 07 Aug 2024 14:15
 Operator : JC/MD
 Sample : P3462-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-3

Quant Time: Aug 07 14:43:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	146997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	276006	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	273229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115369	57.171	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 114.340%		
35) Dibromofluoromethane	5.391	113	96552	53.501	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 107.000%		
50) Toluene-d8	8.646	98	337070	52.613	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 105.220%		
62) 4-Bromofluorobenzene	11.079	95	131413	57.082	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 114.160%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.001	59	7153	17.821	ug/l #	92
16) Acetone	2.391	43	59775	68.293	ug/l	91
18) Methyl Acetate	2.715	43	3768	1.358	ug/l	94
20) Methylene Chloride	2.794	84	11754	5.623	ug/l #	72
25) 2-Butanone	4.580	43	26641	19.382	ug/l	99
37) Ethyl Acetate	4.745	43	3865	1.263	ug/l #	93
40) Benzene	6.037	78	75756	9.976	ug/l	96
43) Isopropyl Acetate	6.342	43	166883	36.860	ug/l	97
52) Toluene	8.720	92	70504	15.306	ug/l	94
67) Ethyl Benzene	10.195	91	20995	2.153	ug/l	96
68) m/p-Xylenes	10.299	106	36946	9.923	ug/l	94
69) o-Xylene	10.640	106	28900	7.749	ug/l	98
70) Styrene	10.658	104	7895	1.316	ug/l #	44
80) 1,3,5-Trimethylbenzene	11.451	105	12526	1.729	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	33784	4.632	ug/l	87
95) Naphthalene	13.774	128	591499	69.474	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042913.D
Acq On : 07 Aug 2024 14:15
Operator : JC/MD
Sample : P3462-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-3

Quant Time: Aug 07 14:43:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

