

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042915.D
 Acq On : 07 Aug 2024 15:01
 Operator : JC/MD
 Sample : P3374-18ME
 Misc : 7.76g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-863-K2-SO-12.75-13.25-072024ME

Quant Time: Aug 08 02:02:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	138660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	257840	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241784	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108038	56.757	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.520%
35) Dibromofluoromethane	5.385	113	88786	52.664	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.320%
50) Toluene-d8	8.647	98	302574	50.556	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.079	95	110410	51.338	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.680%
Target Compounds						
16) Acetone	2.416	43	3029	3.669	ug/l	94
18) Methyl Acetate	2.721	43	12814	4.894	ug/l	100
21) trans-1,2-Dichloroethene	3.081	96	2060	1.340	ug/l #	65
27) cis-1,2-Dichloroethene	4.483	96	269191	147.622	ug/l	82
44) Trichloroethene	7.129	130	9723	5.455	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042915.D
Acq On : 07 Aug 2024 15:01
Operator : JC/MD
Sample : P3374-18ME
Misc : 7.76g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-863-K2-SO-12.75-13.25-072024ME

Quant Time: Aug 08 02:02:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

