

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042916.D
 Acq On : 07 Aug 2024 15:25
 Operator : JC/MD
 Sample : P3374-22ME
 Misc : 7.54g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-863-K2-SO-12-12.5-072024ME

Quant Time: Aug 08 02:02:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	146455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	269090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260216	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	111641	55.528	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.060%
35) Dibromofluoromethane	5.385	113	93127	52.930	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.860%
50) Toluene-d8	8.647	98	318885	51.054	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.100%
62) 4-Bromofluorobenzene	11.079	95	126334	56.286	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.580%
Target Compounds						
16) Acetone	2.416	43	2728	3.128	ug/l #	75
18) Methyl Acetate	2.721	43	13009	4.704	ug/l	99
21) trans-1,2-Dichloroethene	3.075	96	2093	1.289	ug/l	83
27) cis-1,2-Dichloroethene	4.477	96	370399	192.312	ug/l	81
44) Trichloroethene	7.123	130	14488	7.788	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042916.D
Acq On : 07 Aug 2024 15:25
Operator : JC/MD
Sample : P3374-22ME
Misc : 7.54g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-863-K2-SO-12-12.5-072024ME

Quant Time: Aug 08 02:02:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

