

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042926.D
 Acq On : 07 Aug 2024 19:18
 Operator : JC/MD
 Sample : P3464-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOPSOIL

Quant Time: Aug 08 02:06:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	153195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	283466	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99642	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115308	54.829	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.660%	
35) Dibromofluoromethane	5.391	113	99079	53.457	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.920%	
50) Toluene-d8	8.646	98	334809	50.885	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.760%	
62) 4-Bromofluorobenzene	11.079	95	113936	48.188	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.380%	
Target Compounds						
20) Methylene Chloride	2.788	84	6784	2.701	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042926.D
Acq On : 07 Aug 2024 19:18
Operator : JC/MD
Sample : P3464-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOPSOIL

Quant Time: Aug 08 02:06:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

