

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042933.D
 Acq On : 08 Aug 2024 00:18
 Operator : JC/MD
 Sample : P3430-05DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 925-K1-WPO-0-0.25-073124DL

Quant Time: Aug 08 04:04:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	280287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117215	55.115	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 110.220%			
35) Dibromofluoromethane	5.385	113	101278	55.263	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.520%			
50) Toluene-d8	8.647	98	335577	51.580	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.160%			
62) 4-Bromofluorobenzene	11.079	95	113784	48.669	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.340%			
Target Compounds						
4) Vinyl Chloride	1.374	62	22446	13.003	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.325	101	7635	4.523	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	182441	89.548	ug/l	83
44) Trichloroethene	7.135	130	2430	1.254	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042933.D
Acq On : 08 Aug 2024 00:18
Operator : JC/MD
Sample : P3430-05DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
925-K1-WPO-0-0.25-073124DL

Quant Time: Aug 08 04:04:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

