

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042934.D
 Acq On : 08 Aug 2024 00:41
 Operator : JC/MD
 Sample : P3410-01DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-S-864-9-14-073024DL

Quant Time: Aug 08 04:05:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	288124	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	121569	56.597	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.200%
35) Dibromofluoromethane	5.385	113	103939	55.172	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.340%
50) Toluene-d8	8.647	98	346401	51.795	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.600%
62) 4-Bromofluorobenzene	11.079	95	122980	51.172	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.340%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	2858	1.676	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	143067	69.528	ug/l	83
44) Trichloroethene	7.123	130	29648	14.885	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042934.D
Acq On : 08 Aug 2024 00:41
Operator : JC/MD
Sample : P3410-01DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-S-864-9-14-073024DL

Quant Time: Aug 08 04:05:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

