

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042935.D
 Acq On : 08 Aug 2024 01:05
 Operator : JC/MD
 Sample : P3393-25
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-03-072624

Quant Time: Aug 08 04:05:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	156179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	287449	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	277579	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	121852	56.833	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.660%
35) Dibromofluoromethane	5.385	113	101939	54.238	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.480%
50) Toluene-d8	8.647	98	351138	52.627	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.260%
62) 4-Bromofluorobenzene	11.079	95	126496	52.759	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.520%
Target Compounds						
16) Acetone	2.392	43	4857	5.223	ug/l #	88
20) Methylene Chloride	2.788	84	4378	1.371	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042935.D
Acq On : 08 Aug 2024 01:05
Operator : JC/MD
Sample : P3393-25
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-03-072624

Quant Time: Aug 08 04:05:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

