

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042936.D
 Acq On : 08 Aug 2024 01:28
 Operator : JC/MD
 Sample : P3393-26
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-02-072624

Quant Time: Aug 08 04:06:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	281577	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127478	62.556	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 125.120%#			
35) Dibromofluoromethane	5.385	113	106903	58.065	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 116.140%			
50) Toluene-d8	8.647	98	318768	48.772	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.540%			
62) 4-Bromofluorobenzene	11.079	95	109614	46.671	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.340%			
Target Compounds						
16) Acetone	2.386	43	4418	4.998	ug/l	97
20) Methylene Chloride	2.788	84	5175	1.930	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042936.D
Acq On : 08 Aug 2024 01:28
Operator : JC/MD
Sample : P3393-26
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-02-072624

Quant Time: Aug 08 04:06:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

