

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042927.D
 Acq On : 07 Aug 2024 19:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 02:07:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	52.040	-4.1	103	0.00
3 P	Chloromethane	50.000	45.478	9.0	99	0.00
4 C	Vinyl Chloride	50.000	48.368	3.3#	100	0.00
5 T	Bromomethane	50.000	46.099	7.8	100	0.00
6 T	Chloroethane	50.000	41.420	17.2	83	0.00
7 T	Trichlorofluoromethane	50.000	46.027	7.9	92	0.00
8 T	Diethyl Ether	50.000	48.937	2.1	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.053	1.9	102	0.00
10 T	Methyl Iodide	50.000	59.533	-19.1	119	0.00
11 T	Tert butyl alcohol	250.000	259.387	-3.8	98	0.00
12 CM	1,1-Dichloroethene	50.000	48.200	3.6#	101	0.00
13 T	Acrolein	250.000	224.668	10.1	92	0.00
14 T	Allyl chloride	50.000	48.349	3.3	101	0.00
15 T	Acrylonitrile	250.000	228.730	8.5	95	0.00
16 T	Acetone	250.000	221.355	11.5	95	0.00
17 T	Carbon Disulfide	50.000	46.122	7.8	97	0.00
18 T	Methyl Acetate	50.000	45.795	8.4	100	0.00
19 T	Methyl tert-butyl Ether	50.000	50.080	-0.2	103	0.00
20 T	Methylene Chloride	50.000	48.239	3.5	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.589	2.8	99	0.00
22 T	Diisopropyl ether	50.000	47.328	5.3	97	0.00
23 T	Vinyl Acetate	250.000	243.974	2.4	97	0.00
24 P	1,1-Dichloroethane	50.000	48.587	2.8	100	0.00
25 T	2-Butanone	250.000	230.246	7.9	94	0.00
26 T	2,2-Dichloropropane	50.000	53.946	-7.9	111	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.292	-0.6	102	0.00
28 T	Bromochloromethane	50.000	39.066	21.9	93	0.00
29 T	Tetrahydrofuran	250.000	222.583	11.0	92	0.00
30 C	Chloroform	50.000	49.430	1.1#	102	0.00
31 T	Cyclohexane	50.000	47.716	4.6	97	0.00
32 T	1,1,1-Trichloroethane	50.000	50.720	-1.4	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.931	2.1	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	49.544	0.9	100	0.00
36 T	1,1-Dichloropropene	50.000	48.391	3.2	100	0.00
37 T	Ethyl Acetate	50.000	46.366	7.3	94	0.00
38 T	Carbon Tetrachloride	50.000	51.086	-2.2	104	0.00
39 T	Methylcyclohexane	50.000	51.718	-3.4	99	0.00
40 TM	Benzene	50.000	49.635	0.7	98	0.00
41 T	Methacrylonitrile	50.000	49.454	1.1	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.777	-1.6	102	0.00
43 T	Isopropyl Acetate	50.000	49.064	1.9	96	0.00
44 TM	Trichloroethene	50.000	50.021	-0.0	102	0.00
45 C	1,2-Dichloropropane	50.000	48.276	3.4#	97	0.00
46 T	Dibromomethane	50.000	50.081	-0.2	100	0.00
47 T	Bromodichloromethane	50.000	51.451	-2.9	100	0.00
48 T	Methyl methacrylate	50.000	49.506	1.0	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042927.D
 Acq On : 07 Aug 2024 19:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 02:07:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1009.863	-1.0	98	0.00
50 S	Toluene-d8	50.000	48.626	2.7	99	0.00
51 T	4-Methyl-2-Pentanone	250.000	238.288	4.7	94	0.00
52 CM	Toluene	50.000	50.357	-0.7#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	52.751	-5.5	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.688	-5.4	102	0.00
55 T	1,1,2-Trichloroethane	50.000	49.543	0.9	100	0.00
56 T	Ethyl methacrylate	50.000	50.967	-1.9	99	0.00
57 T	1,3-Dichloropropane	50.000	50.024	-0.0	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	242.468	3.0	92	0.00
59 T	2-Hexanone	250.000	238.508	4.6	95	0.00
60 T	Dibromochloromethane	50.000	51.070	-2.1	102	0.00
61 T	1,2-Dibromoethane	50.000	49.962	0.1	99	0.00
62 S	4-Bromofluorobenzene	50.000	49.873	0.3	104	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	51.209	-2.4	102	0.00
65 PM	Chlorobenzene	50.000	49.176	1.6	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.791	-3.6	103	0.00
67 C	Ethyl Benzene	50.000	50.738	-1.5#	100	0.00
68 T	m/p-Xylenes	100.000	100.759	-0.8	98	0.00
69 T	o-Xylene	50.000	50.392	-0.8	100	0.00
70 T	Styrene	50.000	52.056	-4.1	99	0.00
71 P	Bromoform	50.000	49.928	0.1	102	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	49.408	1.2	101	0.00
74 T	N-ethyl acetate	50.000	48.258	3.5	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.697	8.6	99	0.00
76 T	1,2,3-Trichloropropane	50.000	44.607	10.8	99	0.00
77 T	Bromobenzene	50.000	48.755	2.5	102	0.00
78 T	n-propylbenzene	50.000	49.762	0.5	99	0.00
79 T	2-Chlorotoluene	50.000	48.698	2.6	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.337	-2.7	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.065	-2.1	102	0.00
82 T	4-Chlorotoluene	50.000	49.055	1.9	101	0.00
83 T	tert-Butylbenzene	50.000	49.615	0.8	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.125	-2.3	101	0.00
85 T	sec-Butylbenzene	50.000	51.284	-2.6	102	0.00
86 T	p-Isopropyltoluene	50.000	51.537	-3.1	100	0.00
87 T	1,3-Dichlorobenzene	50.000	48.616	2.8	102	0.00
88 T	1,4-Dichlorobenzene	50.000	47.685	4.6	103	0.00
89 T	n-Butylbenzene	50.000	50.480	-1.0	100	0.00
90 T	Hexachloroethane	50.000	49.026	1.9	102	0.00
91 T	1,2-Dichlorobenzene	50.000	49.738	0.5	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.537	-1.1	104	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.678	-1.4	105	0.00
94 T	Hexachlorobutadiene	50.000	51.108	-2.2	105	0.00
95 T	Naphthalene	50.000	52.615	-5.2	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042927.D
Acq On : 07 Aug 2024 19:41
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 08 02:07:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.982	-6.0	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6