

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042929.D
 Acq On : 07 Aug 2024 21:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 04:01:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	51.470	-2.9	98	0.00
3 P	Chloromethane	50.000	44.498	11.0	93	0.00
4 C	Vinyl Chloride	50.000	48.129	3.7#	95	0.00
5 T	Bromomethane	50.000	48.481	3.0	101	0.00
6 T	Chloroethane	50.000	42.309	15.4	82	0.00
7 T	Trichlorofluoromethane	50.000	46.563	6.9	89	0.00
8 T	Diethyl Ether	50.000	48.167	3.7	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.949	2.1	98	0.00
10 T	Methyl Iodide	50.000	54.859	-9.7	105	0.00
11 T	Tert butyl alcohol	250.000	269.342	-7.7	98	0.00
12 CM	1,1-Dichloroethene	50.000	47.242	5.5#	95	0.00
13 T	Acrolein	250.000	253.990	-1.6	100	0.00
14 T	Allyl chloride	50.000	49.075	1.8	99	0.00
15 T	Acrylonitrile	250.000	230.534	7.8	92	0.00
16 T	Acetone	250.000	222.601	11.0	92	0.00
17 T	Carbon Disulfide	50.000	45.236	9.5	91	0.00
18 T	Methyl Acetate	50.000	46.030	7.9	96	0.00
19 T	Methyl tert-butyl Ether	50.000	49.106	1.8	97	0.00
20 T	Methylene Chloride	50.000	47.743	4.5	97	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.032	3.9	94	0.00
22 T	Diisopropyl ether	50.000	47.411	5.2	93	0.00
23 T	Vinyl Acetate	250.000	242.097	3.2	92	0.00
24 P	1,1-Dichloroethane	50.000	47.846	4.3	94	0.00
25 T	2-Butanone	250.000	235.067	6.0	92	0.00
26 T	2,2-Dichloropropane	50.000	51.606	-3.2	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.353	1.3	96	0.00
28 T	Bromochloromethane	50.000	43.277	13.4	99	0.00
29 T	Tetrahydrofuran	250.000	227.856	8.9	91	0.00
30 C	Chloroform	50.000	48.250	3.5#	95	0.00
31 T	Cyclohexane	50.000	48.143	3.7	94	0.00
32 T	1,1,1-Trichloroethane	50.000	49.454	1.1	96	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.107	1.8	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	50.819	-1.6	98	0.00
36 T	1,1-Dichloropropene	50.000	49.494	1.0	97	0.00
37 T	Ethyl Acetate	50.000	47.299	5.4	91	0.00
38 T	Carbon Tetrachloride	50.000	51.254	-2.5	99	0.00
39 T	Methylcyclohexane	50.000	51.511	-3.0	94	0.00
40 TM	Benzene	50.000	49.408	1.2	93	0.00
41 T	Methacrylonitrile	50.000	50.003	-0.0	92	0.00
42 TM	1,2-Dichloroethane	50.000	50.065	-0.1	96	0.00
43 T	Isopropyl Acetate	50.000	49.344	1.3	93	0.00
44 TM	Trichloroethene	50.000	50.294	-0.6	98	0.00
45 C	1,2-Dichloropropane	50.000	48.302	3.4#	93	0.00
46 T	Dibromomethane	50.000	49.245	1.5	94	0.00
47 T	Bromodichloromethane	50.000	50.331	-0.7	94	0.00
48 T	Methyl methacrylate	50.000	50.408	-0.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042929.D
 Acq On : 07 Aug 2024 21:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 04:01:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1052.331	-5.2	98	0.00
50 S	Toluene-d8	50.000	49.356	1.3	96	0.00
51 T	4-Methyl-2-Pentanone	250.000	243.849	2.5	92	0.00
52 CM	Toluene	50.000	49.976	0.0#	93	0.00
53 T	t-1,3-Dichloropropene	50.000	51.949	-3.9	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.027	-4.1	96	0.00
55 T	1,1,2-Trichloroethane	50.000	49.063	1.9	94	0.00
56 T	Ethyl methacrylate	50.000	51.473	-2.9	96	0.00
57 T	1,3-Dichloropropane	50.000	49.084	1.8	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	262.309	-4.9	95	0.00
59 T	2-Hexanone	250.000	243.625	2.5	92	0.00
60 T	Dibromochloromethane	50.000	50.316	-0.6	96	0.00
61 T	1,2-Dibromoethane	50.000	49.944	0.1	95	0.00
62 S	4-Bromofluorobenzene	50.000	50.161	-0.3	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	50.089	-0.2	94	0.00
65 PM	Chlorobenzene	50.000	48.660	2.7	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.724	-3.4	97	0.00
67 C	Ethyl Benzene	50.000	51.249	-2.5#	95	0.00
68 T	m/p-Xylenes	100.000	102.494	-2.5	94	0.00
69 T	o-Xylene	50.000	50.611	-1.2	95	0.00
70 T	Styrene	50.000	52.464	-4.9	94	0.00
71 P	Bromoform	50.000	49.937	0.1	97	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	49.267	1.5	96	0.00
74 T	N-amyl acetate	50.000	49.060	1.9	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.834	8.3	95	0.00
76 T	1,2,3-Trichloropropane	50.000	45.453	9.1	96	0.00
77 T	Bromobenzene	50.000	48.928	2.1	97	0.00
78 T	n-propylbenzene	50.000	49.867	0.3	94	0.00
79 T	2-Chlorotoluene	50.000	48.489	3.0	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.849	-1.7	98	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.761	-1.5	97	0.00
82 T	4-Chlorotoluene	50.000	49.728	0.5	98	0.00
83 T	tert-Butylbenzene	50.000	49.568	0.9	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.105	-2.2	97	0.00
85 T	sec-Butylbenzene	50.000	51.324	-2.6	98	0.00
86 T	p-Isopropyltoluene	50.000	52.246	-4.5	97	0.00
87 T	1,3-Dichlorobenzene	50.000	48.660	2.7	98	0.00
88 T	1,4-Dichlorobenzene	50.000	47.688	4.6	99	0.00
89 T	n-Butylbenzene	50.000	50.780	-1.6	96	0.00
90 T	Hexachloroethane	50.000	48.126	3.7	96	0.00
91 T	1,2-Dichlorobenzene	50.000	49.891	0.2	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.502	-1.0	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.025	-2.0	101	0.00
94 T	Hexachlorobutadiene	50.000	50.276	-0.6	99	0.00
95 T	Naphthalene	50.000	52.281	-4.6	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042929.D
Acq On : 07 Aug 2024 21:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 08 04:01:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.209	-4.4	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6