

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX080724\
 Data File : VX042938.D
 Acq On : 08 Aug 2024 02:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 04:06:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	50.000	52.045	-4.1	90	0.00
3 P	Chloromethane	50.000	49.397	1.2	93	0.00
4 C	Vinyl Chloride	50.000	51.820	-3.6#	92	0.00
5 T	Bromomethane	50.000	53.726	-7.5	101	0.00
6 T	Chloroethane	50.000	48.575	2.8	85	0.00
7 T	Trichlorofluoromethane	50.000	49.435	1.1	85	0.00
8 T	Diethyl Ether	50.000	52.982	-6.0	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.282	1.4	89	0.00
10 T	Methyl Iodide	50.000	61.600	-23.2	106	0.00
11 T	Tert butyl alcohol	250.000	288.252	-15.3	94	0.00
12 CM	1,1-Dichloroethene	50.000	50.670	-1.3#	92	0.00
13 T	Acrolein	250.000	252.012	-0.8	89	0.00
14 T	Allyl chloride	50.000	53.423	-6.8	97	0.00
15 T	Acrylonitrile	250.000	254.212	-1.7	92	0.00
16 T	Acetone	250.000	243.715	2.5	91	0.00
17 T	Carbon Disulfide	50.000	48.319	3.4	88	0.00
18 T	Methyl Acetate	50.000	49.357	1.3	93	0.00
19 T	Methyl tert-butyl Ether	50.000	53.047	-6.1	95	0.00
20 T	Methylene Chloride	50.000	52.921	-5.8	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.815	-3.6	92	0.00
22 T	Diisopropyl ether	50.000	50.809	-1.6	90	0.00
23 T	Vinyl Acetate	250.000	264.604	-5.8	91	0.00
24 P	1,1-Dichloroethane	50.000	51.653	-3.3	92	0.00
25 T	2-Butanone	250.000	252.252	-0.9	89	0.00
26 T	2,2-Dichloropropane	50.000	45.775	8.5	82	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.051	-6.1	93	0.00
28 T	Bromochloromethane	50.000	49.435	1.1	102	0.00
29 T	Tetrahydrofuran	250.000	246.874	1.3	89	0.00
30 C	Chloroform	50.000	52.297	-4.6#	93	0.00
31 T	Cyclohexane	50.000	48.814	2.4	86	0.00
32 T	1,1,1-Trichloroethane	50.000	51.937	-3.9	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.468	1.1	90	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	48.049	3.9	88	0.00
36 T	1,1-Dichloropropene	50.000	48.852	2.3	91	0.00
37 T	Ethyl Acetate	50.000	48.664	2.7	89	0.00
38 T	Carbon Tetrachloride	50.000	50.361	-0.7	92	0.00
39 T	Methylcyclohexane	50.000	49.148	1.7	85	0.00
40 TM	Benzene	50.000	50.303	-0.6	90	0.00
41 T	Methacrylonitrile	50.000	51.248	-2.5	90	0.00
42 TM	1,2-Dichloroethane	50.000	51.656	-3.3	94	0.00
43 T	Isopropyl Acetate	50.000	51.330	-2.7	91	0.00
44 TM	Trichloroethene	50.000	49.319	1.4	91	0.00
45 C	1,2-Dichloropropane	50.000	49.212	1.6#	89	0.00
46 T	Dibromomethane	50.000	50.922	-1.8	92	0.00
47 T	Bromodichloromethane	50.000	52.560	-5.1	93	0.00
48 T	Methyl methacrylate	50.000	51.734	-3.5	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042938.D
 Acq On : 08 Aug 2024 02:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 04:06:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1022.844	-2.3	90	0.00
50 S	Toluene-d8	50.000	46.419	7.2	85	0.00
51 T	4-Methyl-2-Pentanone	250.000	253.882	-1.6	91	0.00
52 CM	Toluene	50.000	51.275	-2.5#	91	0.00
53 T	t-1,3-Dichloropropene	50.000	52.643	-5.3	91	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.407	-4.8	91	0.00
55 T	1,1,2-Trichloroethane	50.000	51.582	-3.2	94	0.00
56 T	Ethyl methacrylate	50.000	53.557	-7.1	94	0.00
57 T	1,3-Dichloropropane	50.000	51.758	-3.5	93	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	264.731	-5.9	91	0.00
59 T	2-Hexanone	250.000	253.041	-1.2	91	0.00
60 T	Dibromochloromethane	50.000	52.812	-5.6	95	0.00
61 T	1,2-Dibromoethane	50.000	52.317	-4.6	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.845	2.3	92	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	90	0.00
64 T	Tetrachloroethene	50.000	51.036	-2.1	92	0.00
65 PM	Chlorobenzene	50.000	50.500	-1.0	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.569	-5.1	95	0.00
67 C	Ethyl Benzene	50.000	51.954	-3.9#	92	0.00
68 T	m/p-Xylenes	100.000	104.031	-4.0	92	0.00
69 T	o-Xylene	50.000	51.653	-3.3	92	0.00
70 T	Styrene	50.000	54.272	-8.5	93	0.00
71 P	Bromoform	50.000	52.025	-4.0	97	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	49.598	0.8	93	0.00
74 T	N-amyl acetate	50.000	51.371	-2.7	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.532	4.9	94	0.00
76 T	1,2,3-Trichloropropane	50.000	46.763	6.5	95	0.00
77 T	Bromobenzene	50.000	49.361	1.3	94	0.00
78 T	n-propylbenzene	50.000	50.932	-1.9	93	0.00
79 T	2-Chlorotoluene	50.000	49.646	0.7	95	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.151	-2.3	94	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.311	1.4	90	0.00
82 T	4-Chlorotoluene	50.000	50.383	-0.8	95	0.00
83 T	tert-Butylbenzene	50.000	50.187	-0.4	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.414	-2.8	94	0.00
85 T	sec-Butylbenzene	50.000	51.669	-3.3	94	0.00
86 T	p-Isopropyltoluene	50.000	51.979	-4.0	93	0.00
87 T	1,3-Dichlorobenzene	50.000	49.822	0.4	96	0.00
88 T	1,4-Dichlorobenzene	50.000	48.510	3.0	97	0.00
89 T	n-Butylbenzene	50.000	50.667	-1.3	92	0.00
90 T	Hexachloroethane	50.000	49.037	1.9	94	0.00
91 T	1,2-Dichlorobenzene	50.000	51.185	-2.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.650	-1.3	96	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.315	-0.6	96	0.00
94 T	Hexachlorobutadiene	50.000	50.326	-0.7	95	0.00
95 T	Naphthalene	50.000	52.884	-5.8	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042938.D
Acq On : 08 Aug 2024 02:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 08 04:06:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.674	-5.3	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6