



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042928.D
Acq On : 07 Aug 2024 21:36
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Title : SW846 8260
Last Update : Wed Aug 07 02:55:22 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.9	11045	PASS
75	95	30	60	52.7	34461	PASS
95	95	100	100	100.0	65352	PASS
96	95	5	9	6.9	4508	PASS
173	174	0.00	2	1.1	581	PASS
174	95	50	100	79.5	51979	PASS
175	174	5	9	7.2	3754	PASS
176	174	95	101	91.5	47544	FAIL*
177	176	5	9	6.7	3170	PASS