

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080725\
 Data File : VX047253.D
 Acq On : 07 Aug 2025 11:03
 Operator : JC/MD
 Sample : VX0807WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0807WBL01

Quant Time: Aug 08 02:40:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	248214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	424037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	386661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	198090	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	151287	48.929	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.860%
35) Dibromofluoromethane	5.397	113	129469	49.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	8.653	98	378639	44.657	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.320%#
62) 4-Bromofluorobenzene	11.079	95	181050	49.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080725\
Data File : VX047253.D
Acq On : 07 Aug 2025 11:03
Operator : JC/MD
Sample : VX0807WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0807WBL01

Quant Time: Aug 08 02:40:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

