

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080818\
 Data File : VX003927.D
 Acq On : 08 Aug 2018 16:51
 Operator : JC/MD
 Sample : J4379-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-107S

Quant Time: Aug 09 10:57:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	320954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	502896	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	476849	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	240696	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	223704	52.58	ug/l	0.00
Spiked Amount						
			Recovery	=	105.16%	
35) Dibromofluoromethane	5.51	113	190438	49.99	ug/l	0.00
Spiked Amount						
			Recovery	=	99.98%	
50) Toluene-d8	8.72	98	713103	48.13	ug/l	0.00
Spiked Amount						
			Recovery	=	96.26%	
62) 4-Bromofluorobenzene	11.14	95	251010	52.36	ug/l	0.00
Spiked Amount						
			Recovery	=	104.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
65) Chlorobenzene	10.14	112	15406	1.24	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080818\
Data File : VX003927.D
Acq On : 08 Aug 2018 16:51
Operator : JC/MD
Sample : J4379-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-107S

Quant Time: Aug 09 10:57:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

