

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080818\  
 Data File : VX003928.D  
 Acq On : 08 Aug 2018 17:18  
 Operator : JC/MD  
 Sample : J4383-01  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 JTY-GW-05

Quant Time: Aug 09 10:58:01 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X080618W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 07 07:45:48 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min)      |
|-----------------------------|-------|------|----------|-------|---------|---------------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 291912   | 50.00 | ug/l    | 0.00          |
| 34) 1,4-Difluorobenzene     | 6.87  | 114  | 461847   | 50.00 | ug/l    | 0.00          |
| 63) Chlorobenzene-d5        | 10.12 | 117  | 432044   | 50.00 | ug/l    | 0.00          |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 210175   | 50.00 | ug/l    | 0.00          |
| System Monitoring Compounds |       |      |          |       |         |               |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 205001   | 52.98 | ug/l    | 0.00          |
| Spiked Amount 50.000        |       |      | Recovery | =     | 105.96% |               |
| 35) Dibromofluoromethane    | 5.51  | 113  | 177023   | 50.59 | ug/l    | 0.00          |
| Spiked Amount 50.000        |       |      | Recovery | =     | 101.18% |               |
| 50) Toluene-d8              | 8.72  | 98   | 652067   | 47.92 | ug/l    | 0.00          |
| Spiked Amount 50.000        |       |      | Recovery | =     | 95.84%  |               |
| 62) 4-Bromofluorobenzene    | 11.14 | 95   | 223325   | 50.73 | ug/l    | 0.00          |
| Spiked Amount 50.000        |       |      | Recovery | =     | 101.46% |               |
| Target Compounds            |       |      |          |       |         |               |
| 16) Acetone                 | 2.45  | 43   | 16533    | 7.18  | ug/l    | Qvalue<br>100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080818\  
Data File : VX003928.D  
Acq On : 08 Aug 2018 17:18  
Operator : JC/MD  
Sample : J4383-01  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
JTY-GW-05

Quant Time: Aug 09 10:58:01 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X080618W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 07 07:45:48 2018  
Response via : Initial Calibration

