


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080818\
Data File : VX003915.D
Acq On : 08 Aug 2018 10:16
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Title : SW846 8260
Last Update : Tue Aug 07 07:45:48 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	2213	PASS
75	95	30	60	48.7	6366	PASS
95	95	100	100	100.0	13080	PASS
96	95	5	9	6.8	885	PASS
173	174	0.00	2	1.1	144	PASS
174	95	50	100	96.2	12585	PASS
175	174	5	9	7.2	912	PASS
176	174	95	101	95.4	12002	PASS
177	176	5	9	6.4	767	PASS