

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080819\
 Data File : VX011389.D
 Acq On : 08 Aug 2019 14:05
 Operator : JC/SP
 Sample : K4220-06DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-24B_080719DL

Quant Time: Aug 09 04:50:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	144736	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	235529	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	231107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	105131	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	83971	57.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.22%	
35) Dibromofluoromethane	5.49	113	75768	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
50) Toluene-d8	8.71	98	289292	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	11.13	95	103165	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
Target Compounds						
64) Tetrachloroethene	9.34	164	92703	47.234	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080819\
Data File : VX011389.D
Acq On : 08 Aug 2019 14:05
Operator : JC/SP
Sample : K4220-06DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-24B_080719DL

Quant Time: Aug 09 04:50:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

