

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080822\
 Data File : VX030520.D
 Acq On : 08 Aug 2022 18:33
 Operator : JC/MD
 Sample : N4066-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Aug 09 01:07:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 09 00:14:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	285750	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	256799	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	111237	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	64005	34.879	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.760%
7) Chloroethane-d5	1.666	69	53107	41.053	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.100%
11) 1,1-Dichloroethene-d2	2.307	63	94799	27.311	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	54.620%#
21) 2-Butanone-d5	4.471	46	188465	110.845	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.850%
24) Chloroform-d	5.068	84	162052	42.899	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.800%
26) 1,2-Dichloroethane-d4	5.964	65	117224	45.703	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.400%
32) Benzene-d6	5.983	84	293439	39.892	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.780%
36) 1,2-Dichloropropane-d6	7.318	67	102328	42.876	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.760%
41) Toluene-d8	8.653	98	248253	35.803	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	71.600%#
43) trans-1,3-Dichloroprop...	8.952	79	43481	36.622	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.240%
47) 2-Hexanone-d5	9.391	63	123406	101.445	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.440%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	147733	47.121	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.240%
66) 1,2-Dichlorobenzene-d4	12.323	152	91840	42.567	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080822\
Data File : VX030520.D
Acq On : 08 Aug 2022 18:33
Operator : JC/MD
Sample : N4066-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Aug 09 01:07:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 09 00:14:16 2022
Response via : Initial Calibration

