

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080822\
 Data File : VX030519.D
 Acq On : 08 Aug 2022 18:10
 Operator : JC/MD
 Sample : N4066-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 F5E27DL

Quant Time: Aug 09 01:07:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 09 00:14:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	280865	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	256176	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	111637	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	64571	35.800	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.600%
7) Chloroethane-d5	1.660	69	53406	42.003	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.000%
11) 1,1-Dichloroethene-d2	2.306	63	98358	28.829	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.660%#
21) 2-Butanone-d5	4.470	46	168888	101.059	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.060%
24) Chloroform-d	5.068	84	154224	41.537	ug/L	0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.080%
26) 1,2-Dichloroethane-d4	5.964	65	110415	43.797	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.600%
32) Benzene-d6	5.982	84	295018	40.204	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.400%
36) 1,2-Dichloropropane-d6	7.317	67	99102	41.626	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.260%
41) Toluene-d8	8.653	98	259099	37.459	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	74.920%#
43) trans-1,3-Dichloroprop...	8.957	79	43858	37.030	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.060%
47) 2-Hexanone-d5	9.390	63	109934	90.590	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.590%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	135890	43.449	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.900%
66) 1,2-Dichlorobenzene-d4	12.323	152	91491	42.253	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.500%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.373	62	12353	6.364	ug/L	91
20) cis-1,2-Dichloroethene	4.495	96	77067	37.617	ug/L	99
34) Trichloroethene	7.128	95	200567	84.312	ug/L	88
46) Tetrachloroethene	9.280	164	41860	27.843	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080822\
Data File : VX030519.D
Acq On : 08 Aug 2022 18:10
Operator : JC/MD
Sample : N4066-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
F5E27DL

Quant Time: Aug 09 01:07:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 09 00:14:16 2022
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 6.364 ug/L
 RT: 1.373 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX030519.D
 Acq: 08 Aug 2022 18:10




Instrument :
 MSVOA_X
 ClientSampleId :
 F5E27DL


Tgt Ion: 62 Resp: 12353
 Ion Ratio Lower Upper
 62 100
 64 39.4 23.8 44.2



#20
 cis-1,2-Dichloroethene
 Concen: 37.617 ug/L
 RT: 4.495 min Scan# 559
 Delta R.T. 0.006 min
 Lab File: VX030519.D
 Acq: 08 Aug 2022 18:10

Tgt Ion: 96 Resp: 77067
 Ion Ratio Lower Upper
 96 100
 61 135.5 94.0 174.6
 68 0.0 0.0 0.0









#34
Trichloroethene
Concen: 84.312 ug/L
RT: 7.128 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX030519.D
Acq: 08 Aug 2022 18:10

Instrument :
MSVOA_X
ClientSampleId :
F5E27DL

Tgt Ion: 95 Resp: 200567
Ion Ratio Lower Upper
95 100
97 66.6 40.2 74.6
132 87.2 55.3 102.7
130 96.2 58.9 109.5



#46
Tetrachloroethene
Concen: 27.843 ug/L
RT: 9.280 min Scan# 1344
Delta R.T. 0.006 min
Lab File: VX030519.D
Acq: 08 Aug 2022 18:10

Tgt Ion:164 Resp: 41860
Ion Ratio Lower Upper
164 100
129 97.0 72.7 135.1
131 93.8 69.9 129.9
166 120.1 88.3 164.1

