

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
 Data File : VX042950.D
 Acq On : 08 Aug 2024 13:28
 Operator : JC/MD
 Sample : PB162548ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 09 05:08:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 07 04:53:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	4.904	128	42216	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	235344	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	208747	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	107488	27.573	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	91.900%
60) 4-Bromofluorobenzene	11.079	95	88939	26.115	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.067%
63) Toluene-d8	8.647	98	281382	29.842	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.467%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	1239	2.277	ug/l	81
18) Methylene Chloride	2.788	84	7241	2.471	ug/l	96
23) tert-Butyl Alcohol	2.940	59	7483	9.669	ug/l #	100
56) Bromoform	11.079	173	1017	0.478	ug/l #	1
77) t-1,4-Dichloro-2-butene	11.079	75	45704	39.143	ug/l #	11


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
Data File : VX042950.D
Acq On : 08 Aug 2024 13:28
Operator : JC/MD
Sample : PB162548ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Aug 09 05:08:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 07 04:53:59 2024
Response via : Initial Calibration








#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.904 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX042950.D
Acq: 08 Aug 2024 13:28

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:128 Resp: 42216
Ion Ratio Lower Upper
128 100
49 154.7 0.0 400.5
130 128.0 0.0 333.0



#15
Acetone
Concen: 2.277 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.013 min
Lab File: VX042950.D
Acq: 08 Aug 2024 13:28


Tgt Ion: 58 Resp: 1239
Ion Ratio Lower Upper
58 100
43 297.5 269.9 404.9








#18
Methylene Chloride
Concen: 2.471 ug/l
RT: 2.788 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX042950.D
Acq: 08 Aug 2024 13:28

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 84 Resp: 7241




Ion	Ratio	Lower	Upper
84	100		
49	115.4	95.0	142.6
51	39.3	30.1	45.1
86	64.8	55.4	83.2




#23
tert-Butyl Alcohol
Concen: 9.669 ug/l
RT: 2.940 min Scan# 304
Delta R.T. -0.024 min
Lab File: VX042950.D
Acq: 08 Aug 2024 13:28

Tgt Ion: 59 Resp: 7483

Ion	Ratio	Lower	Upper
59	100		
52	0.0	0.1	0.1#





#27

1,2-Dichloroethane-d4

Concen: 27.573 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX042950.D

Acq: 08 Aug 2024 13:28

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 107488

Ion Ratio Lower Upper

65 100

67 51.3 40.6 60.8



#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX042950.D

Acq: 08 Aug 2024 13:28




Tgt Ion: 114 Resp: 235344


Ion Ratio Lower Upper

114 100

63 19.9 16.0 24.0

88 15.0 12.5 18.7





#56

Bromoform

Concen: 0.478 ug/l

RT: 11.079 min Scan# 1

Delta R.T. 0.281 min


Lab File: VX042950.D

Acq: 08 Aug 2024 13:28

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:173 Resp: 1017

Ion Ratio Lower Upper

173 100

175 489.0 37.1 55.7#

254 0.0 6.6 10.0#



#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX042950.D

Acq: 08 Aug 2024 13:28




Tgt Ion:117 Resp: 208747


Ion Ratio Lower Upper

117 100

82 54.9 44.9 67.3

119 31.9 25.4 38.2









#60
4-Bromofluorobenzene
Concen: 26.115 ug/l
RT: 11.079 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX042950.D
Acq: 08 Aug 2024 13:28




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 95 Resp: 88939
Ion Ratio Lower Upper
95 100
174 78.1 65.0 97.6
176 76.5 63.4 95.0



#63
Toluene-d8
Concen: 29.842 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX042950.D
Acq: 08 Aug 2024 13:28


Tgt Ion: 98 Resp: 281382
Ion Ratio Lower Upper
98 100
100 64.7 50.9 76.3





#77
 t-1,4-Dichloro-2-butene
 Concen: 39.143 ug/l
 RT: 11.079 min Scan# 1
 Delta R.T. 0.061 min
 Lab File: VX042950.D
 Acq: 08 Aug 2024 13:28

Instrument :
 MSVOA_X
 ClientSampleId :



Tgt Ion: 75 Resp: 45704
 Ion Ratio Lower Upper
 75 100
 53 0.0 48.5 145.5#
 89 0.0 21.9 65.5#

