

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
 Data File : VX042963.D
 Acq On : 09 Aug 2024 00:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 09 05:46:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	230896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	388392	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	344983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	171900	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	165530	52.222	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.440%	
35) Dibromofluoromethane	5.385	113	130792	51.503	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.000%	
50) Toluene-d8	8.647	98	454584	50.424	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.840%	
62) 4-Bromofluorobenzene	11.079	95	169054	52.183	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.360%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	128041	51.447	ug/l	97
3) Chloromethane	1.294	50	131186	48.378	ug/l	99
4) Vinyl Chloride	1.374	62	131056	50.938	ug/l	99
5) Bromomethane	1.599	94	77237	49.603	ug/l	95
6) Chloroethane	1.666	64	70510	48.105	ug/l	98
7) Trichlorofluoromethane	1.867	101	190294	47.558	ug/l	99
8) Diethyl Ether	2.136	74	75171	51.076	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.319	101	127617	50.728	ug/l	96
10) Methyl Iodide	2.447	142	149133	57.331	ug/l	94
11) Tert butyl alcohol	3.001	59	160224	254.130	ug/l #	93
12) 1,1-Dichloroethene	2.306	96	125688	49.387	ug/l	94
13) Acrolein	2.239	56	174617	280.978	ug/l	100
14) Allyl chloride	2.660	41	198951	51.736	ug/l	94
15) Acrylonitrile	3.068	53	402559	250.160	ug/l	98
16) Acetone	2.392	43	317836	231.180	ug/l	94
17) Carbon Disulfide	2.501	76	286448	46.253	ug/l	99
18) Methyl Acetate	2.709	43	213179	48.896	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	437437	51.067	ug/l	98
20) Methylene Chloride	2.788	84	145892	51.342	ug/l	89
21) trans-1,2-Dichloroethene	3.087	96	128409	50.151	ug/l	94
22) Diisopropyl ether	3.763	45	438015	51.849	ug/l	99
23) Vinyl Acetate	3.721	43	2531322	264.011	ug/l	99
24) 1,1-Dichloroethane	3.605	63	248910	51.231	ug/l	99
25) 2-Butanone	4.568	43	543727	251.832	ug/l	96
26) 2,2-Dichloropropane	4.477	77	142503	47.710	ug/l	95
27) cis-1,2-Dichloroethene	4.483	96	154500	50.881	ug/l	91
28) Bromochloromethane	4.897	49	86114	46.191	ug/l	90
29) Tetrahydrofuran	5.013	42	359145	246.162	ug/l	96
30) Chloroform	5.092	83	256730	50.475	ug/l	98
31) Cyclohexane	5.470	56	201125	50.711	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	225582	50.380	ug/l	98
36) 1,1-Dichloropropene	5.690	75	167663	48.696	ug/l	97
37) Ethyl Acetate	4.721	43	213816	49.634	ug/l	100
38) Carbon Tetrachloride	5.672	117	191253	49.208	ug/l	98
39) Methylcyclohexane	7.379	83	208147	50.855	ug/l	97
40) Benzene	6.037	78	540540	50.585	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
 Data File : VX042963.D
 Acq On : 09 Aug 2024 00:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 09 05:46:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	117761	53.074	ug/l	95
42) 1,2-Dichloroethane	6.086	62	198686	51.118	ug/l	98
43) Isopropyl Acetate	6.342	43	323706	50.810	ug/l	97
44) Trichloroethene	7.123	130	130004	48.418	ug/l	99
45) 1,2-Dichloropropane	7.427	63	135753	50.081	ug/l	98
46) Dibromomethane	7.580	93	98555	49.504	ug/l	92
47) Bromodichloromethane	7.824	83	190883	51.613	ug/l	98
48) Methyl methacrylate	7.696	41	159465	52.200	ug/l	93
49) 1,4-Dioxane	7.671	88	71425	1053.240	ug/l #	77
51) 4-Methyl-2-Pentanone	8.574	43	1096319	256.900	ug/l	98
52) Toluene	8.720	92	330378	50.970	ug/l	95
53) t-1,3-Dichloropropene	8.976	75	177518	51.925	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	196284	51.987	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	133107	50.514	ug/l	95
56) Ethyl methacrylate	9.116	69	221080	52.139	ug/l	94
57) 1,3-Dichloropropane	9.311	76	221482	51.309	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	520447	260.423	ug/l	99
59) 2-Hexanone	9.433	43	851257	257.430	ug/l	98
60) Dibromochloromethane	9.518	129	145965	50.426	ug/l	99
61) 1,2-Dibromoethane	9.610	107	138850	49.874	ug/l	98
64) Tetrachloroethene	9.275	164	122311	50.990	ug/l	97
65) Chlorobenzene	10.079	112	363438	49.645	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	127582	50.542	ug/l	99
67) Ethyl Benzene	10.195	91	627608	50.967	ug/l	99
68) m/p-Xylenes	10.299	106	486741	103.541	ug/l	98
69) o-Xylene	10.640	106	237902	50.519	ug/l	100
70) Styrene	10.652	104	404874	53.465	ug/l	97
71) Bromoform	10.799	173	106203	49.661	ug/l #	98
73) Isopropylbenzene	10.963	105	617769	49.707	ug/l	99
74) N-amyl acetate	10.841	43	288890	51.660	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	215267	47.281	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	233066	60.286	ug/l	90
77) Bromobenzene	11.195	156	153558	48.572	ug/l	91
78) n-propylbenzene	11.305	91	702040	50.854	ug/l	100
79) 2-Chlorotoluene	11.366	91	431805	48.801	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	523023	50.825	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	59077	49.687	ug/l	95
82) 4-Chlorotoluene	11.451	91	502147	50.490	ug/l	97
83) tert-Butylbenzene	11.713	119	514187	48.620	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	533084	51.456	ug/l	99
85) sec-Butylbenzene	11.890	105	640814	51.459	ug/l	100
86) p-Isopropyltoluene	12.006	119	535150	51.724	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	280005	48.939	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	275470	46.973	ug/l	98
89) n-Butylbenzene	12.329	91	457293	51.234	ug/l	98
90) Hexachloroethane	12.536	117	90957	48.060	ug/l	87
91) 1,2-Dichlorobenzene	12.335	146	285380	49.810	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	49362	49.342	ug/l	86
93) 1,2,4-Trichlorobenzene	13.585	180	169249	48.429	ug/l	97
94) Hexachlorobutadiene	13.725	225	71871	48.490	ug/l	98
95) Naphthalene	13.774	128	613742	50.750	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	177541	50.468	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
Data File : VX042963.D
Acq On : 09 Aug 2024 00:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Aug 09 05:46:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 230896
Ion Ratio Lower Upper
168 100
99 60.4 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 51.447 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 85 Resp: 128041
Ion Ratio Lower Upper
85 100
87 35.0 16.7 50.1









#3
Chloromethane
Concen: 48.378 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 131186
Ion Ratio Lower Upper
50 100
52 31.2 25.5 38.3



#4
Vinyl Chloride
Concen: 50.938 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 62 Resp: 131056
Ion Ratio Lower Upper
62 100
64 32.1 25.4 38.0









#5
Bromomethane
Concen: 49.603 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 77237
Ion Ratio Lower Upper
94 100
96 97.8 74.6 111.8



#6
Chloroethane
Concen: 48.105 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 64 Resp: 70510
Ion Ratio Lower Upper
64 100
66 30.2 25.1 37.7









#7
Trichlorofluoromethane
Concen: 47.558 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 190294
Ion Ratio Lower Upper
101 100
103 63.7 51.4 77.2



#8
Diethyl Ether
Concen: 51.076 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 74 Resp: 75171
Ion Ratio Lower Upper
74 100
45 100.8 53.3 159.8








#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.728 ug/l
RT: 2.319 min Scan# 201
Delta R.T. 0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:101 Resp: 127617
Ion Ratio Lower Upper
101 100
85 44.1 37.0 55.6
151 77.0 58.6 87.8




#10
Methyl Iodide
Concen: 57.331 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35



Tgt Ion:142 Resp: 149133
Ion Ratio Lower Upper
142 100
127 47.0 41.6 62.4
141 14.4 11.6 17.4





#11

Tert butyl alcohol

Concen: 254.130 ug/l

RT: 3.001 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 160224

Ion Ratio Lower Upper

59 100

57 10.0 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 49.387 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35




Tgt Ion: 96 Resp: 125688


Ion Ratio Lower Upper

96 100

61 169.8 144.6 216.8

98 65.1 50.4 75.6





#13

Acrolein

Concen: 280.978 ug/l

RT: 2.239 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 174617

Ion Ratio Lower Upper

56 100

55 71.7 57.1 85.7



#14

Allyl chloride

Concen: 51.736 ug/l

RT: 2.660 min Scan# 258

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35




Tgt Ion: 41 Resp: 198951


Ion Ratio Lower Upper

41 100

39 64.9 55.8 83.6


76 34.5 24.3 36.5








#15
Acrylonitrile
Concen: 250.160 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 402559
Ion Ratio Lower Upper
53 100
52 83.4 66.0 99.0
51 35.5 29.8 44.6



#16
Acetone
Concen: 231.180 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 43 Resp: 317836
Ion Ratio Lower Upper
43 100
58 30.6 21.9 32.9








#17
Carbon Disulfide
Concen: 46.253 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 286448
Ion Ratio Lower Upper
76 100
78 9.3 7.0 10.6



#18
Methyl Acetate
Concen: 48.896 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 43 Resp: 213179
Ion Ratio Lower Upper
43 100
74 23.2 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 51.067 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 437437
Ion Ratio Lower Upper
73 100
57 22.3 17.3 25.9



#20
Methylene Chloride
Concen: 51.342 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 84 Resp: 145892
Ion Ratio Lower Upper
84 100
49 112.7 103.6 155.4
51 35.4 33.8 50.6
86 65.7 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 50.151 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 128409
Ion Ratio Lower Upper
96 100
61 138.8 119.2 178.8
98 62.0 48.7 73.1




#22
Diisopropyl ether
Concen: 51.849 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35



Tgt Ion: 45 Resp: 438015
Ion Ratio Lower Upper
45 100
43 87.0 69.7 104.5
87 27.7 20.8 31.2
59 11.9 9.0 13.4









#23
 Vinyl Acetate
 Concen: 264.011 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX042963.D
 Acq: 09 Aug 2024 00:35




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 2531322
 Ion Ratio Lower Upper
 43 100
 86 10.3 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 51.231 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX042963.D
 Acq: 09 Aug 2024 00:35

Tgt Ion: 63 Resp: 248910
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.3 9.8
 100 4.2 2.1 6.5









#25
2-Butanone
Concen: 251.832 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 543727
Ion Ratio Lower Upper
43 100
72 26.2 19.4 29.2



#26
2,2-Dichloropropane
Concen: 47.710 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 77 Resp: 142503
Ion Ratio Lower Upper
77 100
97 24.0 10.8 32.3








#27
 cis-1,2-Dichloroethene
 Concen: 50.881 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX042963.D
 Acq: 09 Aug 2024 00:35

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion: 96 Resp: 154500
 Ion Ratio Lower Upper
 96 100
 61 139.6 0.0 312.2
 98 64.2 0.0 129.0



#28
 Bromochloromethane
 Concen: 46.191 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.006 min
 Lab File: VX042963.D
 Acq: 09 Aug 2024 00:35

Tgt Ion: 49 Resp: 86114
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 3.8
 130 80.4 57.5 86.3









#29
Tetrahydrofuran
Concen: 246.162 ug/l
RT: 5.013 min Scan# 644
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 42 Resp: 359145
Ion Ratio Lower Upper
42 100
72 45.1 34.2 51.4
71 43.1 32.2 48.2



#30
Chloroform
Concen: 50.475 ug/l
RT: 5.092 min Scan# 657
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 83 Resp: 256730
Ion Ratio Lower Upper
83 100
85 64.2 50.3 75.5








#31
Cyclohexane
Concen: 50.711 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 201125
Ion Ratio Lower Upper
56 100
69 29.1 23.4 35.2
84 86.2 65.9 98.9



#32
1,1,1-Trichloroethane
Concen: 50.380 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 97 Resp: 225582
Ion Ratio Lower Upper
97 100
99 63.9 51.1 76.7
61 43.0 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 52.222 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 165530

Ion Ratio Lower Upper

65 100

67 51.9 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35




Tgt Ion: 114 Resp: 388392


Ion Ratio Lower Upper

114 100

63 20.5 0.0 45.0

88 16.1 0.0 36.6









#35
Dibromofluoromethane
Concen: 51.503 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 130792
Ion Ratio Lower Upper
113 100
111 102.7 82.9 124.3
192 18.7 13.3 19.9



#36
1,1-Dichloropropene
Concen: 48.696 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 75 Resp: 167663
Ion Ratio Lower Upper
75 100
110 35.9 17.0 50.9
77 29.8 24.7 37.1









#37
Ethyl Acetate
Concen: 49.634 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 213816
Ion Ratio Lower Upper
43 100
61 13.5 10.6 16.0
70 10.6 8.3 12.5



#38
Carbon Tetrachloride
Concen: 49.208 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 117 Resp: 191253
Ion Ratio Lower Upper
117 100
119 95.4 78.2 117.2
121 30.5 24.4 36.6









#39
Methylcyclohexane
Concen: 50.855 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 208147
Ion Ratio Lower Upper
83 100
55 76.7 64.6 97.0
98 46.9 36.7 55.1



#40
Benzene
Concen: 50.585 ug/l
RT: 6.037 min Scan# 812
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 78 Resp: 540540
Ion Ratio Lower Upper
78 100
77 22.0 19.3 28.9








#41
Methacrylonitrile
Concen: 53.074 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 117761
Ion Ratio Lower Upper
41 100
39 53.8 45.2 67.8
67 71.6 52.6 78.8
52 30.3 22.7 34.1



#42
1,2-Dichloroethane
Concen: 51.118 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 62 Resp: 198686
Ion Ratio Lower Upper
62 100
98 8.8 0.0 16.0





#43

Isopropyl Acetate

Concen: 50.810 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.006 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 323706

Ion Ratio Lower Upper

43 100

61 20.3 15.2 22.8

87 13.8 9.8 14.6



#44

Trichloroethene

Concen: 48.418 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX042963.D




Acq: 09 Aug 2024 00:35


Tgt Ion:130 Resp: 130004

Ion Ratio Lower Upper

130 100

95 103.1 0.0 204.4









#45
1,2-Dichloropropane
Concen: 50.081 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 135753
Ion Ratio Lower Upper
63 100
65 31.2 24.3 36.5



#46
Dibromomethane
Concen: 49.504 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 93 Resp: 98555
Ion Ratio Lower Upper
93 100
95 85.4 66.6 100.0
174 98.3 69.4 104.0








#47
Bromodichloromethane
Concen: 51.613 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 190883
Ion Ratio Lower Upper
83 100
85 64.1 49.8 74.6
127 9.1 7.1 10.7



#48
Methyl methacrylate
Concen: 52.200 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 41 Resp: 159465
Ion Ratio Lower Upper
41 100
69 83.5 59.2 88.8
39 57.8 47.7 71.5









#49
1,4-Dioxane
Concen: 1053.240 ug/l
RT: 7.671 min Scan# 1080
Delta R.T. 0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 88 Resp: 71425
Ion Ratio Lower Upper
88 100
43 30.1 0.0 0.0#
58 55.8 60.1 90.1#



#50
Toluene-d8
Concen: 50.424 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 98 Resp: 454584
Ion Ratio Lower Upper
98 100
100 64.6 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 256.900 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 1096319
Ion Ratio Lower Upper
43 100
58 39.0 30.4 45.6



#52
Toluene
Concen: 50.970 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 92 Resp: 330378
Ion Ratio Lower Upper
92 100
91 174.1 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 51.925 ug/l
 RT: 8.976 min Scan# 1194
 Delta R.T. -0.006 min
 Lab File: VX042963.D
 Acq: 09 Aug 2024 00:35




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 177518
 Ion Ratio Lower Upper
 75 100
 77 31.6 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 51.987 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX042963.D
 Acq: 09 Aug 2024 00:35

Tgt Ion: 75 Resp: 196284
 Ion Ratio Lower Upper
 75 100
 77 31.8 25.3 37.9
 39 47.9 46.7 70.1





#55

1,1,2-Trichloroethane

Concen: 50.514 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 97 Resp: 133107





Ion Ratio Lower Upper

97 100

83 86.8 65.8 98.8

85 55.7 41.6 62.4

99 62.4 47.8 71.8



#56

Ethyl methacrylate

Concen: 52.139 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35




Tgt Ion: 69 Resp: 221080


Ion Ratio Lower Upper

69 100

41 66.0 56.8 85.2

39 35.5 32.0 48.0





#57

1,3-Dichloropropane

Concen: 51.309 ug/l

RT: 9.311 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 221482

Ion Ratio Lower Upper

76 100

78 33.2 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 260.423 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX042963.D




Acq: 09 Aug 2024 00:35


Tgt Ion: 63 Resp: 520447

Ion Ratio Lower Upper

63 100

106 27.7 22.5 33.7









#59
2-Hexanone
Concen: 257.430 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 851257
Ion Ratio Lower Upper
43 100
58 53.7 26.2 78.5



#60
Dibromochloromethane
Concen: 50.426 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 129 Resp: 145965
Ion Ratio Lower Upper
129 100
127 78.3 38.8 116.4









#61
1,2-Dibromoethane
Concen: 49.874 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 138850
Ion Ratio Lower Upper
107 100
109 95.8 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 52.183 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 95 Resp: 169054
Ion Ratio Lower Upper
95 100
174 79.2 0.0 135.6
176 76.7 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 344983
Ion Ratio Lower Upper
117 100
82 56.4 46.2 69.4
119 33.3 25.5 38.3




#64
Tetrachloroethene
Concen: 50.990 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35



Tgt Ion:164 Resp: 122311
Ion Ratio Lower Upper
164 100
166 128.3 102.7 154.1
129 94.0 80.4 120.6
131 91.6 76.5 114.7





#65

Chlorobenzene

Concen: 49.645 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:112 Resp: 363438

Ion Ratio Lower Upper

112 100

114 32.7 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 50.542 ug/l

RT: 10.165 min Scan# 1489

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35




Tgt Ion:131 Resp: 127582


Ion Ratio Lower Upper

131 100

133 95.5 47.1 141.4


119 64.5 32.2 96.6








#67
Ethyl Benzene
Concen: 50.967 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 91 Resp: 627608
Ion Ratio Lower Upper
91 100
106 31.9 24.9 37.3



#68
m/p-Xylenes
Concen: 103.541 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 106 Resp: 486741
Ion Ratio Lower Upper
106 100
91 203.6 165.5 248.3









#69
o-Xylene
Concen: 50.519 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 237902
Ion Ratio Lower Upper
106 100
91 218.0 108.9 326.8



#70
Styrene
Concen: 53.465 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion:104 Resp: 404874
Ion Ratio Lower Upper
104 100
78 51.8 43.6 65.4
103 55.5 43.6 65.4





#71

Bromoform

Concen: 49.661 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:173 Resp: 106203

Ion Ratio Lower Upper

173 100

175 46.9 24.1 72.4

254 0.1 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35




Tgt Ion:152 Resp: 171900


Ion Ratio Lower Upper

152 100

115 86.0 43.3 129.9

150 171.9 0.0 345.0









#73
Isopropylbenzene
Concen: 49.707 ug/l
RT: 10.963 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 617769
Ion Ratio Lower Upper
105 100
120 26.2 12.9 38.6



#74
N-ethyl acetate
Concen: 51.660 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 43 Resp: 288890
Ion Ratio Lower Upper
43 100
70 41.7 32.9 49.3
55 24.4 19.3 28.9
61 24.0 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 47.281 ug/l

RT: 11.213 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 215267

Ion Ratio Lower Upper

83 100

131 10.3 5.3 15.8

85 64.8 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 60.286 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX042963.D




Acq: 09 Aug 2024 00:35


Tgt Ion: 75 Resp: 233066

Ion Ratio Lower Upper

75 100

77 32.5 19.1 57.5





#77

Bromobenzene

Concen: 48.572 ug/l

RT: 11.195 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 153558

Ion Ratio Lower Upper

156 100

77 157.0 88.4 265.3

158 96.6 48.0 144.0



#78

n-propylbenzene

Concen: 50.854 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX042963.D




Acq: 09 Aug 2024 00:35


Tgt Ion: 91 Resp: 702040

Ion Ratio Lower Upper

91 100

120 22.8 11.5 34.4





#79

2-Chlorotoluene

Concen: 48.801 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 431805

Ion Ratio Lower Upper

91 100

126 34.1 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 50.825 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX042963.D




Acq: 09 Aug 2024 00:35


Tgt Ion: 105 Resp: 523023

Ion Ratio Lower Upper

105 100

120 47.4 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 49.687 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 59077
Ion Ratio Lower Upper
75 100
53 105.6 90.5 135.7
89 45.3 36.9 55.3



#82
4-Chlorotoluene
Concen: 50.490 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 91 Resp: 502147
Ion Ratio Lower Upper
91 100
126 29.9 14.2 42.6





#83

tert-Butylbenzene

Concen: 48.620 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min


Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:119 Resp: 514187

Ion Ratio Lower Upper

119 100

91 61.3 31.2 93.6

134 22.9 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 51.456 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX042963.D




Acq: 09 Aug 2024 00:35


Tgt Ion:105 Resp: 533084

Ion Ratio Lower Upper

105 100


120 43.2 21.2 63.6








#85
sec-Butylbenzene
Concen: 51.459 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:105 Resp: 640814
Ion Ratio Lower Upper
105 100
134 19.4 9.8 29.4



#86
p-Isopropyltoluene
Concen: 51.724 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion:119 Resp: 535150
Ion Ratio Lower Upper
119 100
134 26.1 12.8 38.3
91 25.1 13.1 39.3








#87
1,3-Dichlorobenzene
Concen: 48.939 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:146 Resp: 280005
Ion Ratio Lower Upper
146 100
111 40.1 22.1 66.1
148 63.3 31.8 95.3




#88
1,4-Dichlorobenzene
Concen: 46.973 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35




Tgt Ion:146 Resp: 275470
Ion Ratio Lower Upper
146 100
111 40.7 21.3 63.7
148 64.6 31.8 95.4








#89
n-Butylbenzene
Concen: 51.234 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 91 Resp: 457293
Ion Ratio Lower Upper
91 100
92 52.2 27.4 82.0
134 25.2 12.6 37.8



#90
Hexachloroethane
Concen: 48.060 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35


Tgt Ion: 117 Resp: 90957
Ion Ratio Lower Upper
117 100
201 76.4 32.9 98.6








#91
1,2-Dichlorobenzene
Concen: 49.810 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 285380
Ion Ratio Lower Upper
146 100
111 42.7 22.4 67.0
148 63.6 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.342 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX042963.D
Acq: 09 Aug 2024 00:35

Tgt Ion: 75 Resp: 49362
Ion Ratio Lower Upper
75 100
155 81.7 36.1 108.5
157 104.4 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 48.429 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 169249

Ion Ratio Lower Upper

180 100

182 94.7 48.3 144.8

145 31.4 18.0 54.0



#94

Hexachlorobutadiene

Concen: 48.490 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35




Tgt Ion:225 Resp: 71871


Ion Ratio Lower Upper

225 100

223 63.0 31.1 93.2

227 62.4 32.5 97.5





#95

Naphthalene

Concen: 50.750 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX042963.D





Acq: 09 Aug 2024 00:35

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	128	Resp:	613742
Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.3	15.5
129	10.7	8.8	13.2



#96

1,2,3-Trichlorobenzene

Concen: 50.468 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX042963.D

Acq: 09 Aug 2024 00:35

Tgt Ion:	180	Resp:	177541
Ion	Ratio	Lower	Upper
180	100		
182	94.1	46.9	140.6
145	33.7	17.8	53.5

