

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
 Data File : VX042947.D
 Acq On : 08 Aug 2024 12:18
 Operator : JC/MD
 Sample : PB162548ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB162548ZHE#01

Quant Time: Aug 09 04:42:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	299951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	121278	56.582	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.160%
35) Dibromofluoromethane	5.385	113	104406	53.235	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.480%
50) Toluene-d8	8.647	98	362587	52.078	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.160%
62) 4-Bromofluorobenzene	11.079	95	130638	52.215	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.440%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.782	84	7308	2.907	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
Data File : VX042947.D
Acq On : 08 Aug 2024 12:18
Operator : JC/MD
Sample : PB162548ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB162548ZHE#01

Quant Time: Aug 09 04:42:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

