

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
 Data File : VX042975.D
 Acq On : 09 Aug 2024 06:26
 Operator : JC/MD
 Sample : P3515-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT189D2-HYD-20240806

Quant Time: Aug 09 08:35:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	151371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	287298	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	269551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	119968	57.732	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.460%
35) Dibromofluoromethane	5.385	113	100437	53.467	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.940%
50) Toluene-d8	8.647	98	357228	53.568	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.140%
62) 4-Bromofluorobenzene	11.079	95	121629	50.755	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.520%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	3388	3.759	ug/l #	78
60) Dibromochloromethane	9.525	129	2694	1.258	ug/l	93
71) Bromoform	10.805	173	995	1.419	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
Data File : VX042975.D
Acq On : 09 Aug 2024 06:26
Operator : JC/MD
Sample : P3515-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT189D2-HYD-20240806

Quant Time: Aug 09 08:35:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

