

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
 Data File : VX042979.D
 Acq On : 09 Aug 2024 07:59
 Operator : JC/MD
 Sample : P3469-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MLS-15-513-528

Quant Time: Aug 09 08:56:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	152850	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	288119	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	278058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	125352	59.739	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	119.480%#
35) Dibromofluoromethane	5.391	113	104077	55.247	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.500%
50) Toluene-d8	8.647	98	358600	53.620	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.240%
62) 4-Bromofluorobenzene	11.079	95	121492	50.554	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.100%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	30276	33.266	ug/l	# 80
52) Toluene	8.720	92	268710	55.884	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
Data File : VX042979.D
Acq On : 09 Aug 2024 07:59
Operator : JC/MD
Sample : P3469-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MLS-15-513-528

Quant Time: Aug 09 08:56:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

