

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
 Data File : VX042980.D
 Acq On : 09 Aug 2024 08:22
 Operator : JC/MD
 Sample : P3469-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MLS-15-625-640

Quant Time: Aug 09 08:57:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148718	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	284201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	123607	60.544	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 121.080%#	
35) Dibromofluoromethane	5.385	113	102535	55.178	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.360%	
50) Toluene-d8	8.647	98	352904	53.496	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.000%	
62) 4-Bromofluorobenzene	11.079	95	118572	50.019	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.040%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	11610	13.111	ug/l	96
52) Toluene	8.720	92	126180	26.604	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080824\
Data File : VX042980.D
Acq On : 09 Aug 2024 08:22
Operator : JC/MD
Sample : P3469-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MLS-15-625-640

Quant Time: Aug 09 08:57:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

