

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080824\  
 Data File : VX042941.D  
 Acq On : 08 Aug 2024 09:47  
 Operator : JC/MD  
 Sample : VX0808WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0808WBS01

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 08/08/2024  
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 14:50:30 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X080624W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Wed Aug 07 04:53:59 2024  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Bromochloromethane        | 4.903          | 128  | 42764      | 30.000   | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 6.757          | 114  | 249168     | 30.000   | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 10.055         | 117  | 225401     | 30.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 27) 1,2-Dichloroethane-d4    | 5.952          | 65   | 110937     | 28.093   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = | 93.633%  |        |          |
| 60) 4-Bromofluorobenzene     | 11.079         | 95   | 109826     | 29.866   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = | 99.567%  |        |          |
| 63) Toluene-d8               | 8.647          | 98   | 307632     | 30.215   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = | 100.733% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 46225      | 18.381   | ug/l   | 100      |
| 3) Chloromethane             | 1.300          | 50   | 40370      | 17.490   | ug/l   | 100      |
| 4) Vinyl Chloride            | 1.373          | 62   | 43459      | 17.636   | ug/l   | 94       |
| 5) Bromomethane              | 1.599          | 94   | 26073      | 17.361   | ug/l   | 96       |
| 6) Chloroethane              | 1.666          | 64   | 25065      | 17.103   | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 1.873          | 101  | 72831      | 18.535   | ug/l   | 97       |
| 8) Diethyl Ether             | 2.136          | 74   | 27539      | 17.832   | ug/l   | 97       |
| 9) 1,1,2-Trichlorotrifluo... | 2.318          | 101  | 48557      | 18.124   | ug/l   | 98       |
| 10) 1,1-Dichloroethene       | 2.312          | 96   | 44220      | 17.279   | ug/l   | 94       |
| 11) Methyl Iodide            | 2.446          | 142  | 47130      | 20.013   | ug/l   | 93       |
| 12) Methyl Acetate           | 2.709          | 43   | 110534     | 19.759   | ug/l   | 97       |
| 13) Acrolein                 | 2.239          | 56   | 14074      | 87.008   | ug/l   | 93       |
| 14) Acrylonitrile            | 3.068          | 53   | 145273     | 85.430   | ug/l   | 97       |
| 15) Acetone                  | 2.386          | 58   | 46364      | 84.097   | ug/l   | 98       |
| 16) Carbon Disulfide         | 2.501          | 76   | 79794      | 14.984   | ug/l # | 95       |
| 17) Allyl chloride           | 2.660          | 41   | 77105      | 18.619   | ug/l   | 99       |
| 18) Methylene Chloride       | 2.788          | 84   | 48025      | 16.181   | ug/l   | 95       |
| 19) trans-1,2-Dichloroethene | 3.087          | 96   | 43939      | 16.861   | ug/l   | 98       |
| 20) Diisopropyl ether        | 3.763          | 45   | 163261     | 18.049   | ug/l   | 95       |
| 21) 1,1-Dichloroethane       | 3.605          | 63   | 93151      | 18.109   | ug/l   | 99       |
| 22) cis-1,2-Dichloroethene   | 4.483          | 96   | 57360      | 17.540   | ug/l   | 98       |
| 23) tert-Butyl Alcohol       | 2.989          | 59   | 67506      | 86.106   | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 3.117          | 73   | 170911     | 18.452   | ug/l   | 99       |
| 25) Chloroform               | 5.092          | 83   | 100881     | 18.125   | ug/l   | 96       |
| 26) Cyclohexane              | 5.464          | 56   | 66482      | 16.384   | ug/l # | 100      |
| 29) 1,1-Dichloropropene      | 5.690          | 75   | 57883      | 18.026   | ug/l   | 99       |
| 30) 2-Butanone               | 4.562          | 43   | 215936     | 91.319   | ug/l   | 99       |
| 31) 2,2-Dichloropropane      | 4.471          | 77   | 78111      | 20.818   | ug/l   | 100      |
| 32) 1,1,1-Trichloroethane    | 5.379          | 97   | 85763      | 19.292   | ug/l   | 98       |
| 33) Carbon Tetrachloride     | 5.672          | 117  | 70381      | 18.508   | ug/l   | 95       |
| 34) Benzene                  | 6.037          | 78   | 193851     | 18.857   | ug/l   | 97       |
| 35) Methacrylonitrile        | 4.922          | 41   | 41379      | 18.522   | ug/l   | 89       |
| 36) 1,2-Dichloroethane       | 6.086          | 62   | 72926      | 19.222   | ug/l   | 98       |
| 37) Trichloroethene          | 7.122          | 130  | 47934      | 18.418   | ug/l   | 98       |
| 38) Methylcyclohexane        | 7.372          | 83   | 72696      | 18.268   | ug/l   | 99       |
| 39) 1,2-Dichloropropane      | 7.427          | 63   | 48363      | 18.454   | ug/l   | 92       |
| 40) Dibromomethane           | 7.580          | 93   | 36401      | 18.396   | ug/l   | 96       |
| 41) Bromodichloromethane     | 7.824          | 83   | 73111      | 18.633   | ug/l   | 99       |
| 42) Vinyl Acetate            | 3.721          | 43   | 634354     | 91.725   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080824\  
 Data File : VX042941.D  
 Acq On : 08 Aug 2024 09:47  
 Operator : JC/MD  
 Sample : VX0808WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0808WBS01

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 08/08/2024  
 Supervised By : Mahesh Dadoda 08/09/2024

Quant Time: Aug 08 14:50:30 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X080624W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Wed Aug 07 04:53:59 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| 43) Ethyl Acetate             | 4.720  | 43   | 69047    | 17.406  | ug/l  | 100      |
| 44) Isopropyl Acetate         | 6.342  | 43   | 118653   | 18.585  | ug/l  | 99       |
| 45) 1,4-Dioxane               | 7.665  | 88   | 26748m   | 400.208 | ug/l  |          |
| 46) Methyl methacrylate       | 7.696  | 41   | 58305    | 18.377  | ug/l  | 96       |
| 47) n-amyl Acetate            | 10.841 | 43   | 94053    | 17.222  | ug/l  | 98       |
| 48) t-1,3-Dichloropropene     | 8.976  | 75   | 69738    | 18.540  | ug/l  | 96       |
| 49) cis-1,3-Dichloropropene   | 8.366  | 75   | 76282    | 18.726  | ug/l  | 95       |
| 50) 1,1,2-Trichloroethane     | 9.153  | 97   | 51619    | 19.127  | ug/l  | 96       |
| 51) Ethyl methacrylate        | 9.116  | 69   | 80005    | 18.118  | ug/l  | 98       |
| 52) 1,3-Dichloropropane       | 9.305  | 76   | 84242    | 19.021  | ug/l  | 99       |
| 53) Dibromochloromethane      | 9.518  | 129  | 56042    | 17.844  | ug/l  | 99       |
| 54) 1,2-Dibromoethane         | 9.610  | 107  | 53596    | 19.096  | ug/l  | 97       |
| 55) 2-Chloroethyl vinyl ether | 8.238  | 63   | 201437   | 98.781  | ug/l  | 98       |
| 56) Bromoform                 | 10.799 | 173  | 39006    | 17.326  | ug/l  | 96       |
| 58) 4-Methyl-2-Pentanone      | 8.573  | 43   | 404551   | 95.308  | ug/l  | 99       |
| 59) 2-Hexanone                | 9.433  | 43   | 310058   | 92.417  | ug/l  | 99       |
| 61) Tetrachloroethene         | 9.268  | 164  | 43657    | 19.509  | ug/l  | 96       |
| 62) Toluene                   | 8.714  | 91   | 211850   | 19.190  | ug/l  | 100      |
| 64) Chlorobenzene             | 10.079 | 112  | 137937   | 19.522  | ug/l  | 96       |
| 65) 1,1,1,2-Tetrachloroethane | 10.159 | 131  | 51139    | 19.937  | ug/l  | 99       |
| 66) Ethyl Benzene             | 10.195 | 91   | 236705   | 19.291  | ug/l  | 96       |
| 67) m/p-Xylenes               | 10.299 | 106  | 180056   | 38.812  | ug/l  | 99       |
| 68) o-Xylene                  | 10.640 | 106  | 90977    | 19.893  | ug/l  | 100      |
| 69) Styrene                   | 10.652 | 104  | 151552   | 19.245  | ug/l  | 99       |
| 70) Isopropylbenzene          | 10.963 | 105  | 234391   | 19.474  | ug/l  | 100      |
| 71) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 83728    | 19.287  | ug/l  | 98       |
| 72) 1,2,3-Trichloropropane    | 11.238 | 75   | 68101m   | 19.205  | ug/l  |          |
| 73) Bromobenzene              | 11.195 | 156  | 56526    | 18.483  | ug/l  | 97       |
| 74) n-propylbenzene           | 11.305 | 91   | 266529   | 18.929  | ug/l  | 99       |
| 75) 2-Chlorotoluene           | 11.366 | 91   | 169066   | 19.673  | ug/l  | 99       |
| 76) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 198432   | 19.451  | ug/l  | 98       |
| 77) t-1,4-Dichloro-2-butene   | 11.018 | 75   | 22465    | 17.818  | ug/l  | 97       |
| 78) 4-Chlorotoluene           | 11.451 | 91   | 187988   | 19.073  | ug/l  | 98       |
| 79) tert-butylbenzene         | 11.713 | 119  | 202709   | 19.365  | ug/l  | 98       |
| 80) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 198012   | 19.674  | ug/l  | 99       |
| 81) sec-Butylbenzene          | 11.890 | 105  | 245765   | 19.156  | ug/l  | 100      |
| 82) p-Isopropyltoluene        | 12.006 | 119  | 203657   | 18.955  | ug/l  | 99       |
| 83) 1,3-Dichlorobenzene       | 11.969 | 146  | 106157   | 18.859  | ug/l  | 99       |
| 84) 1,4-Dichlorobenzene       | 12.042 | 146  | 106538   | 18.825  | ug/l  | 99       |
| 85) n-Butylbenzene            | 12.335 | 91   | 168970   | 17.449  | ug/l  | 99       |
| 86) Hexachloroethane          | 12.536 | 117  | 35146    | 18.125  | ug/l  | 99       |
| 87) 1,2-Dichlorobenzene       | 12.335 | 146  | 108495   | 19.348  | ug/l  | 98       |
| 88) 1,2-Dibromo-3-Chloropr... | 12.938 | 75   | 17807    | 18.158  | ug/l  | 98       |
| 89) 1,2,4-Trichlorobenzene    | 13.585 | 180  | 65924    | 18.070  | ug/l  | 98       |
| 90) Hexachlorobutadiene       | 13.725 | 225  | 30183    | 19.145  | ug/l  | 97       |
| 91) Naphthalene               | 13.774 | 128  | 229673   | 18.701  | ug/l  | 100      |
| 92) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 68270    | 18.530  | ug/l  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080824\  
Data File : VX042941.D  
Acq On : 08 Aug 2024 09:47  
Operator : JC/MD  
Sample : VX0808WBS01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 3 Sample Multiplier: 1


Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Manual Integrations  
APPROVED

Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024


Quant Time: Aug 08 14:50:30 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X080624W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Wed Aug 07 04:53:59 2024  
Response via : Initial Calibration





#1  
Bromochloromethane  
Concen: 30.000 ug/l  
RT: 4.903 min Scan# 61  
Delta R.T. 0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:128 Resp: 42764  
Ion Ratio Lower Upper  
128 100  
49 152.7 0.0 400.5  
130 128.7 0.0 333.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#2  
Dichlorodifluoromethane  
Concen: 18.381 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 85 Resp: 46225  
Ion Ratio Lower Upper  
85 100  
87 31.8 16.0 48.0





#3  
Chloromethane  
Concen: 17.490 ug/l  
RT: 1.300 min Scan# 35  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion: 50 Resp: 40370  
Ion Ratio Lower Upper  
50 100  
52 32.5 26.0 39.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#4  
Vinyl Chloride  
Concen: 17.636 ug/l  
RT: 1.373 min Scan# 47  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 62 Resp: 43459  
Ion Ratio Lower Upper  
62 100  
64 29.6 26.3 39.5









#5  
Bromomethane  
Concen: 17.361 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 94 Resp: 2607  
Ion Ratio Lower Upper  
94 100  
96 98.9 76.2 114.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#6  
Chloroethane  
Concen: 17.103 ug/l  
RT: 1.666 min Scan# 95  
Delta R.T. -0.024 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 64 Resp: 25065  
Ion Ratio Lower Upper  
64 100  
66 33.9 25.9 38.9





#7  
Trichlorofluoromethane  
Concen: 18.535 ug/l  
RT: 1.873 min Scan# 101  
Delta R.T. -0.012 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:101 Resp: 7283  
Ion Ratio Lower Upper  
101 100  
103 66.6 51.5 77.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#8  
Diethyl Ether  
Concen: 17.832 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 74 Resp: 27539  
Ion Ratio Lower Upper  
74 100  
45 97.9 20.3 182.3









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 18.124 ug/l  
RT: 2.318 min Scan# 201  
Delta R.T. -0.012 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion:101 Resp: 4855  
Ion Ratio Lower Upper  
101 100  
85 43.3 0.0 88.6  
151 77.2 0.0 157.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#10  
1,1-Dichloroethene  
Concen: 17.279 ug/l  
RT: 2.312 min Scan# 201  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 96 Resp: 44220  
Ion Ratio Lower Upper  
96 100  
61 154.9 129.2 193.8  
98 67.0 48.6 72.8





#11  
Methyl Iodide  
Concen: 20.013 ug/l  
RT: 2.446 min Scan# 211  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:142 Resp: 47130  
Ion Ratio Lower Upper  
142 100  
127 53.0 23.6 71.0  
141 14.8 8.0 23.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#12  
Methyl Acetate  
Concen: 19.759 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 43 Resp: 110534  
Ion Ratio Lower Upper  
43 100  
74 22.6 19.4 29.2





#13  
Acrolein  
Concen: 87.008 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion: 56 Resp: 14074  
Ion Ratio Lower Upper  
56 100  
55 70.1 60.6 90.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#14  
Acrylonitrile  
Concen: 85.430 ug/l  
RT: 3.068 min Scan# 325  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 53 Resp: 145273  
Ion Ratio Lower Upper  
53 100  
52 85.9 41.2 123.6  
51 37.0 17.9 53.8





#15

Acetone

Concen: 84.097 ug/l

RT: 2.386 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion: 58 Resp: 46364

Ion Ratio Lower Upper

58 100





43 332.7 269.9 404.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#16

Carbon Disulfide

Concen: 14.984 ug/l

RT: 2.501 min Scan# 232

Delta R.T. -0.012 min

Lab File: VX042941.D




Acq: 08 Aug 2024 09:47


Tgt Ion: 76 Resp: 79794

Ion Ratio Lower Upper

76 100


78 10.8 7.1 10.7#





#17  
 Allyl chloride  
 Concen: 18.619 ug/l  
 RT: 2.660 min Scan# 210  
 Delta R.T. -0.006 min  
 Lab File: VX042941.D  
 Acq: 08 Aug 2024 09:47




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0808WBS01



Tgt Ion: 41 Resp: 7710  
 Ion Ratio Lower Upper  
 41 100  
 39 70.4 34.6 103.9  
 76 35.9 18.4 55.4




Manual Integrations  
 APPROVED


Reviewed By :John Carlone 08/08/2024  
 Supervised By :Mahesh Dadoda 08/09/2024



#18  
 Methylene Chloride  
 Concen: 16.181 ug/l  
 RT: 2.788 min Scan# 279  
 Delta R.T. -0.006 min  
 Lab File: VX042941.D  
 Acq: 08 Aug 2024 09:47


Tgt Ion: 84 Resp: 48025  
 Ion Ratio Lower Upper  
 84 100  
 49 116.2 95.0 142.6  
 51 35.3 30.1 45.1  
 86 61.5 55.4 83.2





#19  
trans-1,2-Dichloroethene  
Concen: 16.861 ug/l  
RT: 3.087 min Scan# 31  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion: 96 Resp: 43939  
Ion Ratio Lower Upper  
96 100  
61 138.0 109.6 164.4  
98 65.6 50.4 75.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#20  
Diisopropyl ether  
Concen: 18.049 ug/l  
RT: 3.763 min Scan# 439  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 45 Resp: 163261  
Ion Ratio Lower Upper  
45 100  
43 77.3 65.4 98.2  
87 31.3 23.0 34.6  
59 11.1 9.8 14.6









#21  
1,1-Dichloroethane  
Concen: 18.109 ug/l  
RT: 3.605 min Scan# 414  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 63 Resp: 9315  
Ion Ratio Lower Upper  
63 100  
98 5.7 2.9 8.8  
100 4.5 2.1 6.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#22  
cis-1,2-Dichloroethene  
Concen: 17.540 ug/l  
RT: 4.483 min Scan# 557  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 96 Resp: 57360  
Ion Ratio Lower Upper  
96 100  
61 145.0 117.7 176.5  
98 64.7 53.4 80.0





#23

tert-Butyl Alcohol

Concen: 86.106 ug/l

RT: 2.989 min Scan# 312

Delta R.T. 0.025 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion: 59 Resp: 67500

Ion Ratio Lower Upper

59 100





52 0.2 0.1 0.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#24

Methyl tert-Butyl Ether

Concen: 18.452 ug/l

RT: 3.117 min Scan# 333

Delta R.T. 0.000 min

Lab File: VX042941.D




Acq: 08 Aug 2024 09:47


Tgt Ion: 73 Resp: 170911

Ion Ratio Lower Upper

73 100

43 19.8 16.2 24.2





#25

Chloroform

Concen: 18.125 ug/l

RT: 5.092 min Scan# 61

Delta R.T. -0.006 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion: 83 Resp: 10088

Ion Ratio Lower Upper

83 100





85 64.8 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#26

Cyclohexane

Concen: 16.384 ug/l

RT: 5.464 min Scan# 718

Delta R.T. -0.006 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47




Tgt Ion: 56 Resp: 66482


Ion Ratio Lower Upper

56 100

89 0.0 0.0 0.0

84 89.6 71.8 107.8









#27  
1,2-Dichloroethane-d4  
Concen: 28.093 ug/l  
RT: 5.952 min Scan# 799  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 65 Resp: 11093  
Ion Ratio Lower Upper  
65 100  
67 51.7 40.6 60.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#28  
1,4-Difluorobenzene  
Concen: 30.000 ug/l  
RT: 6.757 min Scan# 930  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion:114 Resp: 249168  
Ion Ratio Lower Upper  
114 100  
63 19.8 16.0 24.0  
88 15.8 12.5 18.7





#29  
1,1-Dichloropropene  
Concen: 18.026 ug/l  
RT: 5.690 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01




Tgt Ion: 75 Resp: 5788  
Ion Ratio Lower Upper  
75 100  
110 37.1 0.0 72.4  
77 31.1 0.0 63.4

Manual Integrations  
APPROVED



Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024




#30  
2-Butanone  
Concen: 91.319 ug/l  
RT: 4.562 min Scan# 570  
Delta R.T. 0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47



Tgt Ion: 43 Resp: 215936  
Ion Ratio Lower Upper  
43 100  
57 8.0 6.6 10.0  
72 26.6 21.0 31.4









#31  
2,2-Dichloropropane  
Concen: 20.818 ug/l  
RT: 4.471 min Scan# 511  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 77 Resp: 7811  
Ion Ratio Lower Upper  
77 100  
97 22.9 18.5 27.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#32  
1,1,1-Trichloroethane  
Concen: 19.292 ug/l  
RT: 5.379 min Scan# 704  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 97 Resp: 85763  
Ion Ratio Lower Upper  
97 100  
99 63.5 52.2 78.2  
61 41.5 34.3 51.5





#33  
Carbon Tetrachloride  
Concen: 18.508 ug/l  
RT: 5.672 min Scan# 753  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:117 Resp: 7038  
Ion Ratio Lower Upper  
117 100  
119 98.5 75.7 113.5  
121 33.1 23.3 34.9




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#34  
Benzene  
Concen: 18.857 ug/l  
RT: 6.037 min Scan# 812  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 78 Resp: 193851  
Ion Ratio Lower Upper  
78 100  
77 23.9 19.4 29.2  
51 14.4 13.5 20.3





#35  
Methacrylonitrile  
Concen: 18.522 ug/l  
RT: 4.922 min Scan# 61  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



| Tgt Ion:  | 41   | Resp: | 41379 |
|-----------|------|-------|-------|
| Ion Ratio | 100  | Lower | Upper |
| 41        | 100  |       |       |
| 39        | 46.2 | 28.1  | 84.2  |
| 67        | 77.0 | 34.4  | 103.1 |
| 52        | 30.4 | 13.3  | 39.8  |




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#36  
1,2-Dichloroethane  
Concen: 19.222 ug/l  
RT: 6.086 min Scan# 820  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


| Tgt Ion:  | 62  | Resp: | 72926 |
|-----------|-----|-------|-------|
| Ion Ratio | 100 | Lower | Upper |
| 62        | 100 |       |       |
| 98        | 9.2 | 6.8   | 10.2  |





#37  
Trichloroethene  
Concen: 18.418 ug/l  
RT: 7.122 min Scan# 991  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:130 Resp: 47934  
Ion Ratio Lower Upper  
130 100  
95 100.0 81.4 122.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#38  
Methylcyclohexane  
Concen: 18.268 ug/l  
RT: 7.372 min Scan# 1031  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 83 Resp: 72696  
Ion Ratio Lower Upper  
83 100  
55 75.4 38.0 113.9  
98 46.0 23.4 70.2





#39

1,2-Dichloropropane

Concen: 18.454 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion: 63 Resp: 4836

Ion Ratio Lower Upper

63 100





65 28.8 26.6 39.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#40

Dibromomethane

Concen: 18.396 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47




Tgt Ion: 93 Resp: 36401


Ion Ratio Lower Upper

93 100

95 84.3 0.0 161.8


174 100.5 0.0 194.0





#41  
Bromodichloromethane  
Concen: 18.633 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01




Tgt Ion: 83 Resp: 7311  
Ion Ratio Lower Upper  
83 100  
85 63.0 50.8 76.2  
127 9.1 6.9 10.3

Manual Integrations  
APPROVED



Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024




#42  
Vinyl Acetate  
Concen: 91.725 ug/l  
RT: 3.721 min Scan# 432  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47



Tgt Ion: 43 Resp: 634354  
Ion Ratio Lower Upper  
43 100  
86 9.2 7.2 10.8









#43  
Ethyl Acetate  
Concen: 17.406 ug/l  
RT: 4.720 min Scan# 596  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 43 Resp: 6904  
Ion Ratio Lower Upper  
43 100  
45 15.6 12.5 18.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#44  
Isopropyl Acetate  
Concen: 18.585 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 43 Resp: 118653  
Ion Ratio Lower Upper  
43 100  
61 19.7 15.4 23.0





#45

1,4-Dioxane

Concen: 400.208 ug/l m

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion: 88 Resp: 2674





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 26.3  | 25.8  | 38.6  |
| 58  | 54.4  | 55.5  | 83.3  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#46

Methyl methacrylate

Concen: 18.377 ug/l

RT: 7.696 min Scan# 1084




Delta R.T. 0.000 min


Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Tgt Ion: 41 Resp: 58305


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 41  | 100   |       |       |
| 69  | 83.8  | 42.4  | 127.1 |
| 39  | 53.0  | 29.9  | 89.7  |





#47  
n-amy1 Acetate  
Concen: 17.222 ug/l  
RT: 10.841 min Scan# 1600  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion: 43 Resp: 9405  
Ion Ratio Lower Upper  
43 100  
55 25.7 19.6 29.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#48  
t-1,3-Dichloropropene  
Concen: 18.540 ug/l  
RT: 8.976 min Scan# 1294  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 75 Resp: 69738  
Ion Ratio Lower Upper  
75 100  
77 29.5 25.6 38.4









#49  
 cis-1,3-Dichloropropene  
 Concen: 18.726 ug/l  
 RT: 8.366 min Scan# 1194  
 Delta R.T. 0.000 min  
 Lab File: VX042941.D  
 Acq: 08 Aug 2024 09:47

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0808WBS01

Tgt Ion: 75 Resp: 7628  
 Ion Ratio Lower Upper  
 75 100  
 77 33.2 26.8 40.2  
 39 52.6 37.5 56.3




Manual Integrations  
 APPROVED


Reviewed By :John Carlone 08/08/2024  
 Supervised By :Mahesh Dadoda 08/09/2024



#50  
 1,1,2-Trichloroethane  
 Concen: 19.127 ug/l  
 RT: 9.153 min Scan# 1323  
 Delta R.T. 0.000 min  
 Lab File: VX042941.D  
 Acq: 08 Aug 2024 09:47

Tgt Ion: 97 Resp: 51619  
 Ion Ratio Lower Upper  
 97 100  
 83 80.2 68.2 102.2  
 85 53.8 45.4 68.0  
 99 60.5 49.0 73.4









#51  
Ethyl methacrylate  
Concen: 18.118 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 69 Resp: 8000  
Ion Ratio Lower Upper  
69 100  
41 67.0 33.1 99.2  
39 35.4 18.8 56.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#52  
1,3-Dichloropropane  
Concen: 19.021 ug/l  
RT: 9.305 min Scan# 1348  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 76 Resp: 84242  
Ion Ratio Lower Upper  
76 100  
78 32.8 0.0 64.8









#53  
Dibromochloromethane  
Concen: 17.844 ug/l  
RT: 9.518 min Scan# 1383  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion:129 Resp: 5604  
Ion Ratio Lower Upper  
129 100  
127 78.0 38.5 115.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#54  
1,2-Dibromoethane  
Concen: 19.096 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion:107 Resp: 53596  
Ion Ratio Lower Upper  
107 100  
109 95.4 74.2 111.4





#55

2-Chloroethyl vinyl ether

Concen: 98.781 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01





|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 63    | Resp: | 20143 |
| Ion Ratio | Lower | Upper |       |
| 63        | 100   |       |       |
| 65        | 32.3  | 26.3  | 39.5  |
| 106       | 28.8  | 21.8  | 32.6  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#56

Bromoform

Concen: 17.326 ug/l




RT: 10.799 min Scan# 1593


Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 173   | Resp: | 39006 |
| Ion Ratio | Lower | Upper |       |
| 173       | 100   |       |       |
| 175       | 49.3  | 37.1  | 55.7  |
| 254       | 8.9   | 6.6   | 10.0  |





#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion:117 Resp: 22540





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 82  | 56.2  | 44.9  | 67.3  |
| 119 | 32.3  | 25.4  | 38.2  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#58

4-Methyl-2-Pentanone

Concen: 95.308 ug/l

RT: 8.573 min Scan# 1228




Delta R.T. 0.000 min


Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Tgt Ion: 43 Resp: 404551


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 58  | 39.7  | 31.4  | 47.0  |
| 85  | 18.7  | 14.6  | 21.8  |





#59  
2-Hexanone  
Concen: 92.417 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion: 43 Resp: 31058  
Ion Ratio Lower Upper  
43 100  
58 54.0 43.5 65.3  
57 19.3 15.0 22.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#60  
4-Bromofluorobenzene  
Concen: 29.866 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 95 Resp: 109826  
Ion Ratio Lower Upper  
95 100  
174 80.6 65.0 97.6  
176 78.2 63.4 95.0





#61  
Tetrachloroethene  
Concen: 19.509 ug/l  
RT: 9.268 min Scan# 11  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:164 Resp: 4365  
Ion Ratio Lower Upper  
164 100  
166 125.8 98.6 147.8  
129 98.8 77.4 116.0  
131 100.6 74.4 111.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#62  
Toluene  
Concen: 19.190 ug/l  
RT: 8.714 min Scan# 1251  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 91 Resp: 211850  
Ion Ratio Lower Upper  
91 100  
92 56.9 45.7 68.5









#63  
Toluene-d8  
Concen: 30.215 ug/l  
RT: 8.647 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 98 Resp: 307632  
Ion Ratio Lower Upper  
98 100  
100 62.7 50.9 76.3




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#64  
Chlorobenzene  
Concen: 19.522 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion:112 Resp: 137937  
Ion Ratio Lower Upper  
112 100  
114 30.7 26.3 39.5









#65  
1,1,1,2-Tetrachloroethane  
Concen: 19.937 ug/l  
RT: 10.159 min Scan# 1488  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion:131 Resp: 51139  
Ion Ratio Lower Upper  
131 100  
133 94.9 0.0 188.2  
119 64.1 0.0 131.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#66  
Ethyl Benzene  
Concen: 19.291 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 91 Resp: 236705  
Ion Ratio Lower Upper  
91 100  
106 32.2 23.8 35.8





#67  
m/p-Xylenes  
Concen: 38.812 ug/l  
RT: 10.299 min Scan# 1511  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01




Tgt Ion:106 Resp: 180050  
Ion Ratio Lower Upper  
106 100  
91 205.3 165.3 247.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024




#68  
o-Xylene  
Concen: 19.893 ug/l  
RT: 10.640 min Scan# 1567  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47



Tgt Ion:106 Resp: 90977  
Ion Ratio Lower Upper  
106 100  
91 215.5 107.5 322.4





#69

Styrene

Concen: 19.245 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01





Tgt Ion:104 Resp: 151552  
 Ion Ratio Lower Upper  
 104 100  
 78 50.2 41.1 61.7  
 103 55.4 44.2 66.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#70

Isopropylbenzene

Concen: 19.474 ug/l




RT: 10.963 min Scan# 1620


Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47


Tgt Ion:105 Resp: 234391  
 Ion Ratio Lower Upper  
 105 100  
 120 26.4 18.4 34.2





#71  
1,1,2,2-Tetrachloroethane  
Concen: 19.287 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01




Tgt Ion: 83 Resp: 83728  
Ion Ratio Lower Upper  
83 100  
131 10.8 4.9 14.7  
85 65.4 32.0 96.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024




#72  
1,2,3-Trichloropropane  
Concen: 19.205 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47



Tgt Ion: 75 Resp: 68101  
Ion Ratio Lower Upper  
75 100  
77 44.8 0.0 89.6





#73

Bromobenzene

Concen: 18.483 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01





|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 156   | Resp: | 56520 |
| Ion Ratio | Lower | Upper |       |
| 156       | 100   |       |       |
| 77        | 161.5 | 0.0   | 317.2 |
| 158       | 101.2 | 0.0   | 195.2 |

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#74

n-propylbenzene

Concen: 18.929 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 91    | Resp: | 266529 |
| Ion Ratio | Lower | Upper |        |
| 91        | 100   |       |        |
| 120       | 22.5  | 0.0   | 45.8   |





#75  
2-Chlorotoluene  
Concen: 19.673 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion: 91 Resp: 169060  
Ion Ratio Lower Upper  
91 100  
126 32.5 0.0 66.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#76  
1,3,5-Trimethylbenzene  
Concen: 19.451 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion:105 Resp: 198432  
Ion Ratio Lower Upper  
105 100  
120 46.1 0.0 94.8









#77  
t-1,4-Dichloro-2-butene  
Concen: 17.818 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01

Tgt Ion: 75 Resp: 2246  
Ion Ratio Lower Upper  
75 100  
53 101.2 48.5 145.5  
89 44.6 21.9 65.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#78  
4-Chlorotoluene  
Concen: 19.073 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47


Tgt Ion: 91 Resp: 187988  
Ion Ratio Lower Upper  
91 100  
126 30.5 0.0 59.4





#79  
tert-butylbenzene  
Concen: 19.365 ug/l  
RT: 11.713 min Scan# 1743  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01




Tgt Ion:119 Resp: 202709  
Ion Ratio Lower Upper  
119 100  
91 60.5 0.0 118.6  
134 22.9 0.0 44.4

Manual Integrations  
APPROVED



Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024




#80  
1,2,4-Trimethylbenzene  
Concen: 19.674 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Tgt Ion:105 Resp: 198012  
Ion Ratio Lower Upper  
105 100  
120 43.4 0.0 87.8





#81  
 sec-Butylbenzene  
 Concen: 19.156 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. 0.000 min  
 Lab File: VX042941.D  
 Acq: 08 Aug 2024 09:47




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0808WBS01




Tgt Ion:105 Resp: 24576  
 Ion Ratio Lower Upper  
 105 100  
 134 19.9 0.0 40.0

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 08/08/2024  
 Supervised By :Mahesh Dadoda 08/09/2024




#82  
 p-Isopropyltoluene  
 Concen: 18.955 ug/l  
 RT: 12.006 min Scan# 1791  
 Delta R.T. 0.000 min  
 Lab File: VX042941.D  
 Acq: 08 Aug 2024 09:47



Tgt Ion:119 Resp: 203657  
 Ion Ratio Lower Upper  
 119 100  
 134 25.6 0.0 51.8  
 91 25.0 0.0 48.4





#83

1,3-Dichlorobenzene

Concen: 18.859 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion:146 Resp: 10615





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 111 | 42.3  | 20.3  | 61.0  |
| 148 | 63.8  | 31.7  | 95.1  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#84

1,4-Dichlorobenzene

Concen: 18.825 ug/l

RT: 12.042 min Scan# 1797




Delta R.T. 0.000 min


Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Tgt Ion:146 Resp: 106538

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 146 | 100   |       |       |
| 111 | 39.5  | 19.4  | 58.4  |
| 148 | 63.8  | 31.4  | 94.2  |





#85

n-Butylbenzene

Concen: 17.449 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01

Tgt Ion: 91 Resp: 168970

Ion Ratio Lower Upper

91 100

92 51.9 0.0 106.4





134 25.1 0.0 50.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#86

Hexachloroethane

Concen: 18.125 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX042941.D




Acq: 08 Aug 2024 09:47


Tgt Ion:117 Resp: 35146

Ion Ratio Lower Upper

117 100


201 74.6 37.5 112.7





#87  
1,2-Dichlorobenzene  
Concen: 19.348 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:146 Resp: 10849  
Ion Ratio Lower Upper  
146 100  
111 41.8 21.0 63.0  
148 64.5 30.8 92.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#88  
1,2-Dibromo-3-Chloropropane  
Concen: 18.158 ug/l  
RT: 12.938 min Scan# 1944  
Delta R.T. -0.006 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion: 75 Resp: 17807  
Ion Ratio Lower Upper  
75 100  
155 87.6 68.7 103.1  
157 105.4 86.7 130.1





#89

1,2,4-Trichlorobenzene

Concen: 18.070 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47

Instrument :

MSVOA\_X

ClientSampleId :

VX0808WBS01





|             |             |
|-------------|-------------|
| Tgt Ion:180 | Resp: 65924 |
| Ion Ratio   | Lower Upper |
| 180 100     |             |
| 182 94.5    | 73.1 109.7  |
| 145 30.9    | 24.6 37.0   |

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/08/2024

Supervised By :Mahesh Dadoda 08/09/2024



#90

Hexachlorobutadiene

Concen: 19.145 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. 0.000 min

Lab File: VX042941.D

Acq: 08 Aug 2024 09:47


|             |             |
|-------------|-------------|
| Tgt Ion:225 | Resp: 30183 |
| Ion Ratio   | Lower Upper |
| 225 100     |             |
| 223 63.3    | 0.0 135.0   |
| 227 63.9    | 0.0 130.4   |





#91  
Naphthalene  
Concen: 18.701 ug/l  
RT: 13.774 min Scan# 2081  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0808WBS01



Tgt Ion:128 Resp: 22967  
Ion Ratio Lower Upper  
128 100  
127 13.0 10.2 15.4  
129 10.8 8.8 13.2

Manual Integrations  
APPROVED

Reviewed By :John Carlone 08/08/2024  
Supervised By :Mahesh Dadoda 08/09/2024



#92  
1,2,3-Trichlorobenzene  
Concen: 18.530 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. 0.000 min  
Lab File: VX042941.D  
Acq: 08 Aug 2024 09:47

Tgt Ion:180 Resp: 68270  
Ion Ratio Lower Upper  
180 100  
182 95.5 0.0 185.8  
145 33.6 0.0 65.2

