

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080825\
 Data File : VX047275.D
 Acq On : 08 Aug 2025 16:33
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 11 02:03:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	301091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	581043	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	548141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	271668	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	212711	56.713	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.420%
35) Dibromofluoromethane	5.398	113	179174	49.943	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.880%
50) Toluene-d8	8.653	98	619284	53.303	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.600%
62) 4-Bromofluorobenzene	11.079	95	264177	52.764	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080825\
Data File : VX047275.D
Acq On : 08 Aug 2025 16:33
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 11 02:03:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

