

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080825\
 Data File : VX047281.D
 Acq On : 08 Aug 2025 18:39
 Operator : JC/MD
 Sample : Q2812-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Aug 11 02:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	193286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	336798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	299963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	153513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106856	44.380	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.760%
35) Dibromofluoromethane	5.403	113	94420	45.405	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.820%
50) Toluene-d8	8.653	98	268901	39.929	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	79.860%#
62) 4-Bromofluorobenzene	11.079	95	131377	45.269	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080825\
Data File : VX047281.D
Acq On : 08 Aug 2025 18:39
Operator : JC/MD
Sample : Q2812-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Aug 11 02:06:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

