

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080918\
 Data File : VX003949.D
 Acq On : 09 Aug 2018 14:39
 Operator : JC/MD
 Sample : J4387-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0557

Quant Time: Aug 10 01:03:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-5.67
34) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/l	-6.86
63) Chlorobenzene-d5	9.95	117	442840	50.00	ug/l	-0.17
72) 1,4-Dichlorobenzene-d4	12.04	152	246217	50.00	ug/l	-0.04

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount				50.000		
Recovery						0.00%
35) Dibromofluoromethane	5.58	113	14824	0.00	ug/l	0.08
Spiked Amount				50.000		
Recovery						0.00%
50) Toluene-d8	8.70	98	102	0.00	ug/l	-0.02
Spiked Amount				50.000		
Recovery						0.00%
62) 4-Bromofluorobenzene	11.06	95	234693	0.00	ug/l	-0.08
Spiked Amount				50.000		
Recovery						0.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
67) Ethyl Benzene	10.22	91	6399	0.34	ug/l	# 67
68) m/p-Xylenes	10.23	106	3250	0.44	ug/l	96
71) Bromoform	11.06	173	1869	3.44	ug/l	# 1
76) 1,2,3-Trichloropropane	11.06	75	111864	20.57	ug/l	# 36
81) trans-1,4-Dichloro-2-buten	11.06	75	111864	55.69	ug/l	# 9
84) 1,2,4-Trimethylbenzene	11.76	105	15390	1.03	ug/l	99
88) 1,4-Dichlorobenzene	11.99	146	604	Below	Cal	87
94) Hexachlorobutadiene	13.78	225	288	0.53	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080918\
Data File : VX003949.D
Acq On : 09 Aug 2018 14:39
Operator : JC/MD
Sample : J4387-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0557

Quant Time: Aug 10 01:03:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

