

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080918\
 Data File : VX003955.D
 Acq On : 09 Aug 2018 17:17
 Operator : JC/MD
 Sample : J4379-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-105S

Quant Time: Aug 10 16:03:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	321674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	501969	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	480122	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	259224	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	218699	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
35) Dibromofluoromethane	5.51	113	189964	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
50) Toluene-d8	8.72	98	709988	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
62) 4-Bromofluorobenzene	11.14	95	253810	53.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.08%	
Target Compounds						
22) Diisopropyl ether	3.87	45	10582	0.86	ug/l	91
65) Chlorobenzene	10.14	112	38398	3.07	ug/l	98
68) m/p-Xylenes	10.36	106	4285	0.54	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080918\
Data File : VX003955.D
Acq On : 09 Aug 2018 17:17
Operator : JC/MD
Sample : J4379-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-105S

Quant Time: Aug 10 16:03:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

