

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080919\
 Data File : VX011406.D
 Acq On : 09 Aug 2019 13:45
 Operator : JC/SP
 Sample : K4279-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-TB-2019-1

Quant Time: Aug 11 00:27:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	202403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	294549	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	119822	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	104284	51.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.32%	
35) Dibromofluoromethane	5.49	113	89534	46.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.12%	
50) Toluene-d8	8.71	98	342791	47.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.44%	
62) 4-Bromofluorobenzene	11.13	95	112507	43.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.82%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080919\
Data File : VX011406.D
Acq On : 09 Aug 2019 13:45
Operator : JC/SP
Sample : K4279-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-TB-2019-1

Quant Time: Aug 11 00:27:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

