

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080919\
 Data File : VX011417.D
 Acq On : 09 Aug 2019 18:02
 Operator : JC/SP
 Sample : K4278-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0691

Quant Time: Aug 11 00:59:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211597	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	312594	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	277616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130613	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	110703	51.95	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.90%	
35) Dibromofluoromethane	5.49	113	92869	45.50	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.00%	
50) Toluene-d8	8.71	98	367728	47.73	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.46%	
62) 4-Bromofluorobenzene	11.13	95	123141	44.77	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	14691	7.277	ug/l	94
21) trans-1,2-Dichloroethene	3.16	96	3945	1.937	ug/l	91
27) cis-1,2-Dichloroethene	4.59	96	58313	24.859	ug/l	94
40) Benzene	6.14	78	31237	3.835	ug/l	97
65) Chlorobenzene	10.13	112	315598	53.020	ug/l	99
67) Ethyl Benzene	10.25	91	3662	0.367	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	1483	0.339	ug/l	96
88) 1,4-Dichlorobenzene	12.10	146	11823	2.620	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080919\
Data File : VX011417.D
Acq On : 09 Aug 2019 18:02
Operator : JC/SP
Sample : K4278-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0691

Quant Time: Aug 11 00:59:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

