


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080919\
Data File : VX011397.D
Acq On : 09 Aug 2019 09:04
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Title : SW846 8260
Last Update : Fri Aug 02 01:55:00 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	5444	PASS
75	95	30	60	51.8	16098	PASS
95	95	100	100	100.0	31058	PASS
96	95	5	9	6.8	2097	PASS
173	174	0.00	2	1.8	526	PASS
174	95	50	100	92.2	28626	PASS
175	174	5	9	8.4	2398	PASS
176	174	95	101	98.7	28248	PASS
177	176	5	9	6.4	1808	PASS