

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
 Data File : VX023606.D
 Acq On : 09 Aug 2021 11:57
 Operator : JC/MD
 Sample : VX0809WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0809WBL02

Quant Time: Aug 10 04:02:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	156801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	273779	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99660	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109951	56.213	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.420%
35) Dibromofluoromethane	5.397	113	85769	50.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.200%
50) Toluene-d8	8.653	98	330437	52.541	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.080%
62) 4-Bromofluorobenzene	11.085	95	111110	51.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
Data File : VX023606.D
Acq On : 09 Aug 2021 11:57
Operator : JC/MD
Sample : VX0809WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0809WBL02

Quant Time: Aug 10 04:02:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

