

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
 Data File : VX023610.D
 Acq On : 09 Aug 2021 13:40
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 10 04:03:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	183719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	305665	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286499	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	122663	53.524	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.040%	
35) Dibromofluoromethane	5.397	113	95468	49.953	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.900%	
50) Toluene-d8	8.653	98	374639	53.355	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.720%	
62) 4-Bromofluorobenzene	11.085	95	137752	57.694	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.380%	
Target Compounds						
						Qvalue
20) Methylene Chloride	2.788	84	1645	0.752	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
Data File : VX023610.D
Acq On : 09 Aug 2021 13:40
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 10 04:03:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

