

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
 Data File : VX023616.D
 Acq On : 09 Aug 2021 16:24
 Operator : JC/MD
 Sample : M3278-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Aug 10 04:03:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	194788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	331945	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	306351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	131709	54.206	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.420%	
35) Dibromofluoromethane	5.391	113	104748	50.470	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.940%	
50) Toluene-d8	8.653	98	403085	52.861	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.720%	
62) 4-Bromofluorobenzene	11.085	95	142787	55.068	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.140%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4104	3.443	ug/l	96
20) Methylene Chloride	2.800	84	972	0.419	ug/l #	85
30) Chloroform	5.105	83	1897	0.491	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
Data File : VX023616.D
Acq On : 09 Aug 2021 16:24
Operator : JC/MD
Sample : M3278-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Aug 10 04:03:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

