

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
 Data File : VX023619.D
 Acq On : 09 Aug 2021 17:35
 Operator : JC/MD
 Sample : M3276-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PRE-MAEC-01

Quant Time: Aug 10 04:04:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	178921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	301172	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	280857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	119445	53.518	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 107.040%		
35) Dibromofluoromethane	5.397	113	95798	50.874	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.740%		
50) Toluene-d8	8.653	98	369461	53.403	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 106.800%		
62) 4-Bromofluorobenzene	11.085	95	129537	55.063	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 110.120%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	3104	6.088	ug/l	97
16) Acetone	2.386	43	22296	20.365	ug/l	95
20) Methylene Chloride	2.794	84	15689	7.367	ug/l #	88
25) 2-Butanone	4.574	43	17246	10.147	ug/l	91
27) cis-1,2-Dichloroethene	4.501	96	2408	1.110	ug/l	96
43) Isopropyl Acetate	6.354	43	24868	4.672	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
Data File : VX023619.D
Acq On : 09 Aug 2021 17:35
Operator : JC/MD
Sample : M3276-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PRE-MAEC-01

Quant Time: Aug 10 04:04:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

