

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
 Data File : VX023620.D
 Acq On : 09 Aug 2021 17:58
 Operator : JC/MD
 Sample : M3276-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PRE-MAEC-02

Quant Time: Aug 10 04:04:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	183025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	313572	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	294220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	123546	54.114	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 108.220%		
35) Dibromofluoromethane	5.397	113	98303	50.139	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.280%		
50) Toluene-d8	8.653	98	380898	52.879	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 105.760%		
62) 4-Bromofluorobenzene	11.085	95	134344	54.848	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 109.700%		
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	3253	6.237	ug/l	98
16) Acetone	2.386	43	22410	20.010	ug/l	94
20) Methylene Chloride	2.794	84	15149	6.954	ug/l	90
25) 2-Butanone	4.581	43	17060	9.812	ug/l	92
27) cis-1,2-Dichloroethene	4.501	96	4020	1.812	ug/l	94
43) Isopropyl Acetate	6.348	43	25697	4.637	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080921\
Data File : VX023620.D
Acq On : 09 Aug 2021 17:58
Operator : JC/MD
Sample : M3276-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PRE-MAEC-02

Quant Time: Aug 10 04:04:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

