

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
 Data File : VX036882.D
 Acq On : 09 Aug 2023 10:54
 Operator : JC/MD
 Sample : VX0809MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0809MBL01

Quant Time: Aug 10 01:13:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	265040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100462	46.980	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.960%
35) Dibromofluoromethane	5.385	113	84077	46.152	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.300%
50) Toluene-d8	8.647	98	317964	49.218	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.440%
62) 4-Bromofluorobenzene	11.079	95	121504	49.851	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.700%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	5586	2.831	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
Data File : VX036882.D
Acq On : 09 Aug 2023 10:54
Operator : JC/MD
Sample : VX0809MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0809MBL01

Quant Time: Aug 10 01:13:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

