

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
 Data File : VX036890.D
 Acq On : 09 Aug 2023 16:20
 Operator : JC/MD
 Sample : 03903-01ME
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-01ME

Quant Time: Aug 10 01:18:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	106545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	194555	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72570	46.033	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.060%
35) Dibromofluoromethane	5.391	113	61229	45.787	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.580%
50) Toluene-d8	8.647	98	235056	49.567	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.140%
62) 4-Bromofluorobenzene	11.079	95	96167	53.750	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.500%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.776	84	2417	1.661	ug/l	90
89) n-Butylbenzene	12.329	91	7580	1.313	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
Data File : VX036890.D
Acq On : 09 Aug 2023 16:20
Operator : JC/MD
Sample : 03903-01ME
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01ME

Quant Time: Aug 10 01:18:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

