

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
 Data File : VX036891.D
 Acq On : 09 Aug 2023 16:44
 Operator : JC/MD
 Sample : 03903-11ME 20X
 Misc : 5.63g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Aug 10 01:19:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226429	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86607	47.035	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.080%
35) Dibromofluoromethane	5.385	113	72410	46.526	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.647	98	275792	49.970	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	109514	52.593	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.180%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	2304	1.356	ug/l	90
39) Methylcyclohexane	7.379	83	3713	1.287	ug/l	95
67) Ethyl Benzene	10.195	91	22554	2.617	ug/l	98
68) m/p-Xylenes	10.305	106	4470	1.390	ug/l	100
73) Isopropylbenzene	10.963	105	13326	1.547	ug/l	100
78) n-propylbenzene	11.305	91	64870	6.510	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	79119	10.977	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	407591	56.543	ug/l	96
89) n-Butylbenzene	12.329	91	18744	2.882	ug/l #	29

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
Data File : VX036891.D
Acq On : 09 Aug 2023 16:44
Operator : JC/MD
Sample : 03903-11ME 20X
Misc : 5.63g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Aug 10 01:19:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

