

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
 Data File : VX036892.D
 Acq On : 09 Aug 2023 17:08
 Operator : JC/MD
 Sample : 03924-01ME 20X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1115ME

Quant Time: Aug 10 01:19:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232157	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87728	46.707	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.420%		
35) Dibromofluoromethane	5.385	113	74056	46.409	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.820%		
50) Toluene-d8	8.647	98	278734	49.257	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.520%		
62) 4-Bromofluorobenzene	11.079	95	108209	50.684	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.360%		
Target Compounds						Qvalue
20) Methylene Chloride	2.788	84	1859	1.072	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
Data File : VX036892.D
Acq On : 09 Aug 2023 17:08
Operator : JC/MD
Sample : 03924-01ME 20X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1115ME

Quant Time: Aug 10 01:19:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

