

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
 Data File : VX036893.D
 Acq On : 09 Aug 2023 17:32
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 10 01:19:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128373	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236939	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104488	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90697	47.749	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.500%
35) Dibromofluoromethane	5.385	113	76162	46.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.540%
50) Toluene-d8	8.647	98	284783	49.310	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.620%
62) 4-Bromofluorobenzene	11.079	95	110733	50.820	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
Data File : VX036893.D
Acq On : 09 Aug 2023 17:32
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 10 01:19:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

