

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
 Data File : VX036896.D
 Acq On : 09 Aug 2023 18:45
 Operator : JC/MD
 Sample : 03957-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15-0809

Quant Time: Aug 10 01:21:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	230914	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90030	48.371	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.740%
35) Dibromofluoromethane	5.385	113	74729	47.083	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.160%
50) Toluene-d8	8.647	98	275363	48.923	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.079	95	102347	48.197	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.400%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	2121	0.401	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
Data File : VX036896.D
Acq On : 09 Aug 2023 18:45
Operator : JC/MD
Sample : 03957-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15-0809

Quant Time: Aug 10 01:21:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

