

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
 Data File : VX036899.D
 Acq On : 09 Aug 2023 19:57
 Operator : JC/MD
 Sample : 03957-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK-1

Quant Time: Aug 10 06:46:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118146	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218749	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83769	47.919	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.840%
35) Dibromofluoromethane	5.385	113	69503	46.226	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.460%
50) Toluene-d8	8.647	98	260990	48.948	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.900%
62) 4-Bromofluorobenzene	11.079	95	100567	49.992	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.980%
Target Compounds						
20) Methylene Chloride	2.794	84	1855	1.150	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080923\
Data File : VX036899.D
Acq On : 09 Aug 2023 19:57
Operator : JC/MD
Sample : 03957-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK-1

Quant Time: Aug 10 06:46:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

