

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
 Data File : VX042987.D
 Acq On : 09 Aug 2024 11:28
 Operator : JC/MD
 Sample : VX0809MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0809MBL01

Quant Time: Aug 09 16:26:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	186171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	299817	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	125810	49.226	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.460%		
35) Dibromofluoromethane	5.385	113	97687	49.831	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.660%		
50) Toluene-d8	8.647	98	363265	52.198	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	104.400%		
62) 4-Bromofluorobenzene	11.079	95	121446	48.563	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.120%		
Target Compounds						
16) Acetone	2.398	43	1434	1.294	ug/l #	74
17) Carbon Disulfide	2.501	76	2642	0.529	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
Data File : VX042987.D
Acq On : 09 Aug 2024 11:28
Operator : JC/MD
Sample : VX0809MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0809MBL01

Quant Time: Aug 09 16:26:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

