

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
 Data File : VX042991.D
 Acq On : 09 Aug 2024 13:06
 Operator : JC/MD
 Sample : P3514-01 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR-05-459-475-080724

Quant Time: Aug 09 16:27:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	192798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	308709	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	129843	49.058	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.120%
35) Dibromofluoromethane	5.385	113	99093	49.093	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	8.647	98	374593	52.276	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.560%
62) 4-Bromofluorobenzene	11.079	95	122988	47.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.520%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1313	1.144	ug/l	95
17) Carbon Disulfide	2.502	76	1916	0.371	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080924\
Data File : VX042991.D
Acq On : 09 Aug 2024 13:06
Operator : JC/MD
Sample : P3514-01 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR-05-459-475-080724

Quant Time: Aug 09 16:27:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

